

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051719\
 Data File : P0056323.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 May 2019 19:58
 Operator : SM/SJ
 Sample : K2884-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MW-00-20190515

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 23:46:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 17 23:42:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.258	3.577	710809	649991	15.315	15.468
2) SA Decachlor...	9.843	8.517	470541	388021	9.797	9.555

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051719\
Data File : P0056323.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 May 2019 19:58
Operator : SM/SJ
Sample : K2884-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MW-00-20190515

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 23:46:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 17 23:42:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.258 min
Delta R.T.: 0.000 min
Response: 710809
Conc: 15.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
MW-00-20190515




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: 0.000 min
Response: 649991
Conc: 15.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.843 min
Delta R.T.: 0.000 min
Response: 470541
Conc: 9.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.517 min
Delta R.T.: -0.002 min
Response: 388021
Conc: 9.55 ng/ml