

Data Path : P:\HPCHEM1\ECD_0\Data\P0051818\
 Data File : P0044781.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 May 2018 22:07
 Operator : SM/UA
 Sample : J2926-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 051518-DBRI-F-D-M

Manual Integrations
 APPROVED

ugo
 5/20/2018 11:12:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 02:13:06 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 15 15:44:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.025	3.293	3275567	4100421	21.375	17.105
2) SA Decachlor...	9.332	8.028	3446734	4737778	16.763m	16.837m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\P0051818\
Data File : P0044781.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 May 2018 22:07
Operator : SM/UA
Sample : J2926-01
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
051518-DBRI-F-D-M


Manual Integrations
APPROVED

ugo
5/20/2018 11:12:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 02:13:06 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: -0.008 min
Response: 3275567
Conc: 21.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
051518-DBRI-F-D-M


Manual Integrations
APPROVED

ugo
5/20/2018 11:12:47 AM




#1 Tetrachloro-m-xylene

R.T.: 3.293 min
Delta R.T.: -0.005 min
Response: 4100421
Conc: 17.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.332 min
Delta R.T.: -0.018 min
Response: 3446734
Conc: 16.76 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.028 min
Delta R.T.: -0.010 min
Response: 4737778
Conc: 16.84 ng/ml m