

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0052019.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 21 03:48:44 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0056371.D 1000 =P0056369.D 750 =P0056370.D
 250 =P0056372.D 50 =P0056373.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.831	2.701	2.780	2.890	3.199	2.880 E4	6.64
2) SA	Decachlorobiphenyl	4.841	4.426	4.615	5.230	6.402	5.103 E4	15.40
3) L1	AR-1016-1	1.470	1.334	1.390	1.588	1.930	1.542 E3	15.36
4) L1	AR-1016-2	2.054	1.900	1.969	2.216	2.663	2.160 E3	14.10
5) L1	AR-1016-3	1.335	1.210	1.262	1.432	1.721	1.392 E3	14.52
6) L1	AR-1016-4	1.117	1.014	1.054	1.178	1.384	1.149 E3	12.62
7) L1	AR-1016-5	1.193	1.068	1.119	1.265	1.464	1.222 E3	12.66
8) L2	AR-1221-1	3.596					3.596 E2	0.00
9) L2	AR-1221-2	2.815					2.815 E2	0.00
10) L2	AR-1221-3	8.841					8.841 E2	0.00
11) L3	AR-1232-1	1.125					1.125 E3	0.00
12) L3	AR-1232-2	6.614					6.614 E2	0.00
13) L3	AR-1232-3	1.421					1.421 E3	0.00
14) L3	AR-1232-4	7.487					7.487 E2	0.00
15) L3	AR-1232-5	5.906					5.906 E2	0.00
16) L4	AR-1242-1	1.171					1.171 E3	0.00
17) L4	AR-1242-2	1.633					1.633 E3	0.00
18) L4	AR-1242-3	1.075					1.075 E3	0.00
19) L4	AR-1242-4	8.886					8.886 E2	0.00
20) L4	AR-1242-5	1.139					1.139 E3	0.00
21) L5	AR-1248-1	9.054					9.054 E2	0.00
22) L5	AR-1248-2	1.317					1.317 E3	0.00
23) L5	AR-1248-3	1.487					1.487 E3	0.00
24) L5	AR-1248-4	1.823					1.823 E3	0.00
25) L5	AR-1248-5	1.756					1.756 E3	0.00
26) L6	AR-1254-1	2.339					2.339 E3	0.00
27) L6	AR-1254-2	3.741					3.741 E3	0.00
28) L6	AR-1254-3	4.188					4.188 E3	0.00
29) L6	AR-1254-4	3.253					3.253 E3	0.00
30) L6	AR-1254-5	3.537					3.537 E3	0.00
31) L7	AR-1260-1	2.359	2.160	2.252	2.581	3.047	2.480 E3	14.26
32) L7	AR-1260-2	2.897	2.685	2.787	3.164	3.853	3.077 E3	15.24
33) L7	AR-1260-3	2.279	2.131	2.202	2.497	2.935	2.409 E3	13.48
34) L7	AR-1260-4	2.652	2.475	2.550	2.826	3.329	2.767 E3	12.32
35) L7	AR-1260-5	5.176	4.930	5.027	5.463	6.366	5.392 E3	10.76
36) L8	AR-1262-1	3.423					3.423 E3	0.00
37) L8	AR-1262-2	5.977					5.977 E3	0.00
38) L8	AR-1262-3	4.125					4.125 E3	0.00
39) L8	AR-1262-4	3.150					3.150 E3	0.00
40) L8	AR-1262-5	2.113					2.113 E3	0.00
41) L9	AR-1268-1	6.643					6.643 E3	0.00
42) L9	AR-1268-2	6.163					6.163 E3	0.00
43) L9	AR-1268-3	5.135					5.135 E3	0.00
44) L9	AR-1268-4	2.241					2.241 E3	0.00
45) L9	AR-1268-5	1.556					1.556 E4	0.00

Signal #2 Calibration Files

500 =P0056371.D 1000 =P0056369.D 750 =P0056370.D
 250 =P0056372.D 50 =P0056373.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.623	2.535	2.561	2.687	3.080	2.697 E4	8.23
2) SA	Decachlorobiphenyl	4.215	3.994	4.093	4.407	5.257	4.393 E4	11.54
3) L1	AR-1016-1	1.310	1.204	1.252	1.416	1.587	1.354 E3	11.26
4) L1	AR-1016-2	1.807	1.678	1.729	1.923	2.233	1.874 E3	11.79
5) L1	AR-1016-3	1.044	0.958	0.996	1.099	1.236	1.066 E3	10.19

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0052019.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 21 03:48:44 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0056371.D 1000 =P0056369.D 750 =P0056370.D
 250 =P0056372.D 50 =P0056373.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	0.840	0.759	0.793	0.894	1.041	0.865	E3	12.80
7)	L1 AR-1016-5	1.135	1.039	1.081	1.213	1.321	1.158	E3	9.68
8)	L2 AR-1221-1	3.174					3.174	E2	0.00
9)	L2 AR-1221-2	2.404					2.404	E2	0.00
10)	L2 AR-1221-3	7.684					7.684	E2	0.00
11)	L3 AR-1232-1	9.843					9.843	E2	0.00
12)	L3 AR-1232-2	1.241					1.241	E3	0.00
13)	L3 AR-1232-3	7.103					7.103	E2	0.00
14)	L3 AR-1232-4	6.191					6.191	E2	0.00
15)	L3 AR-1232-5	7.119					7.119	E2	0.00
16)	L4 AR-1242-1	1.043					1.043	E3	0.00
17)	L4 AR-1242-2	1.420					1.420	E3	0.00
18)	L4 AR-1242-3	8.276					8.276	E2	0.00
19)	L4 AR-1242-4	8.028					8.028	E2	0.00
20)	L4 AR-1242-5	1.107					1.107	E3	0.00
21)	L5 AR-1248-1	8.012					8.012	E2	0.00
22)	L5 AR-1248-2	1.081					1.081	E3	0.00
23)	L5 AR-1248-3	1.125					1.125	E3	0.00
24)	L5 AR-1248-4	1.424					1.424	E3	0.00
25)	L5 AR-1248-5	1.531					1.531	E3	0.00
26)	L6 AR-1254-1	3.009					3.009	E3	0.00
27)	L6 AR-1254-2	2.700					2.700	E3	0.00
28)	L6 AR-1254-3	4.612					4.612	E3	0.00
29)	L6 AR-1254-4	3.026					3.026	E3	0.00
30)	L6 AR-1254-5	4.323					4.323	E3	0.00
31)	L7 AR-1260-1	2.314	2.138	2.210	2.504	3.096	2.453	E3	15.71
32)	L7 AR-1260-2	2.980	2.788	2.865	3.236	4.023	3.178	E3	15.78
33)	L7 AR-1260-3	2.730	2.553	2.629	2.941	3.541	2.879	E3	13.82
34)	L7 AR-1260-4	2.379	2.224	2.277	2.570	3.026	2.495	E3	13.02
35)	L7 AR-1260-5	5.914	5.699	5.748	6.103	7.211	6.135	E3	10.14
36)	L8 AR-1262-1	3.857					3.857	E3	0.00
37)	L8 AR-1262-2	6.632					6.632	E3	0.00
38)	L8 AR-1262-3	2.750					2.750	E3	0.00
39)	L8 AR-1262-4	4.752					4.752	E3	0.00
40)	L8 AR-1262-5	2.132					2.132	E3	0.00
41)	L9 AR-1268-1	7.567					7.567	E3	0.00
42)	L9 AR-1268-2	6.682					6.682	E3	0.00
43)	L9 AR-1268-3	5.669					5.669	E3	0.00
44)	L9 AR-1268-4	2.303					2.303	E3	0.00
45)	L9 AR-1268-5	1.568					1.568	E4	0.00

(#) = Out of Range