

Data Path : P:\HPCHEM1\ECD 0\Data\P0052118\
Data File : P0044889.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 May 2018 15:45
Operator : SM/UA
Sample : J2954-01RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
FK-PS-01-026RE

Manual Integrations
APPROVED

ugo
5/23/2018 10:05:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 16:06:33 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.025	3.290	4492690	5174362	29.318m	21.585m#
2) SA Decachlor...	9.329	8.024	1979599	2608022	9.628	9.269m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO052118\
Data File : PO044889.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 May 2018 15:45
Operator : SM/UA
Sample : J2954-01RE
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
FK-PS-01-026RE


Manual Integrations
APPROVED

ugo
5/23/2018 10:05:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 16:06:33 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: -0.008 min
Response: 4492690
Conc: 29.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
FK-PS-01-026RE


Manual Integrations
APPROVED

ugo
5/23/2018 10:05:23 AM




#1 Tetrachloro-m-xylene

R.T.: 3.290 min
Delta R.T.: -0.008 min
Response: 5174362
Conc: 21.58 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.329 min
Delta R.T.: -0.022 min
Response: 1979599
Conc: 9.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.024 min
Delta R.T.: -0.014 min
Response: 2608022
Conc: 9.27 ng/ml m