

Data Path : P:\HPCHEM1\ECD 0\Data\P0052218\
 Data File : P0044923.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 May 2018 18:13
 Operator : SM/UA
 Sample : J3008-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 051618-JETW-F-U-B

Manual Integrations
 APPROVED

Sohil
 5/24/2018 1:19:23 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 02:07:46 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 15 15:44:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.018	3.289	3145233	4880790	20.525	20.360
2) SA Decachlor...	9.319	8.021	3647484	4442960	17.739m	15.790m

Target Compounds

 (f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO052218\
Data File : PO044923.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 May 2018 18:13
Operator : SM/UA
Sample : J3008-04
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
051618-JETW-F-U-B


Manual Integrations
APPROVED

Sohil
5/24/2018 1:19:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 02:07:46 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.018 min
Delta R.T.: -0.015 min
Response: 3145233
Conc: 20.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
051618-JETW-F-U-B


Manual Integrations
APPROVED

Sohil
5/24/2018 1:19:23 AM




#1 Tetrachloro-m-xylene

R.T.: 3.289 min
Delta R.T.: -0.009 min
Response: 4880790
Conc: 20.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.319 min
Delta R.T.: -0.031 min
Response: 3647484
Conc: 17.74 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.021 min
Delta R.T.: -0.017 min
Response: 4442960
Conc: 15.79 ng/ml m