

Data Path : P:\HPCHEM1\ECD 0\Data\P0052218\
 Data File : P0044934.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 May 2018 20:58
 Operator : SM/UA
 Sample : J3008-13
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 051618-SIDI-F-U-M

Manual Integrations
 APPROVED

Sohil
 5/24/2018 1:20:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 05:31:51 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 15 15:44:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.018	3.289	4094869	5768865	26.722	24.065
2) SA Decachlor...	9.318	8.021	4311097	5695037	20.967m	20.239m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO052218\
Data File : PO044934.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 May 2018 20:58
Operator : SM/UA
Sample : J3008-13
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
051618-SIDI-F-U-M


Manual Integrations
APPROVED

Sohil
5/24/2018 1:20:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 05:31:51 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.018 min
Delta R.T.: -0.014 min
Response: 4094869
Conc: 26.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
051618-SIDI-F-U-M


Manual Integrations
APPROVED

Sohil
5/24/2018 1:20:35 AM




#1 Tetrachloro-m-xylene

R.T.: 3.289 min
Delta R.T.: -0.009 min
Response: 5768865
Conc: 24.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.318 min
Delta R.T.: -0.032 min
Response: 4311097
Conc: 20.97 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.021 min
Delta R.T.: -0.018 min
Response: 5695037
Conc: 20.24 ng/ml m