

Data Path : P:\HPCHEM1\ECD_0\Data\P0052318\
Data File : P0044959.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 May 2018 17:50
Operator : SM/UA
Sample : J3008-14
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
051618-SIDI-F-U-S

Manual Integrations
APPROVED

ugo
5/24/2018 9:13:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 02:26:32 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.017	3.289	4794013	5435334	31.284	22.674 #
2) SA Decachlor...	9.318	8.018	4148220	5406677	20.175	19.215m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\P0052318\
Data File : P0044959.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 May 2018 17:50
Operator : SM/UA
Sample : J3008-14
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
051618-SIDI-F-U-S


Manual Integrations
APPROVED

ugo
5/24/2018 9:13:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 02:26:32 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.017 min
Delta R.T.: -0.016 min
Response: 4794013
Conc: 31.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
051618-SIDI-F-U-S


Manual Integrations
APPROVED

ugo
5/24/2018 9:13:29 PM




#1 Tetrachloro-m-xylene

R.T.: 3.289 min
Delta R.T.: -0.009 min
Response: 5435334
Conc: 22.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.318 min
Delta R.T.: -0.032 min
Response: 4148220
Conc: 20.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.018 min
Delta R.T.: -0.020 min
Response: 5406677
Conc: 19.21 ng/ml m