

Data Path : P:\HPCHEM1\ECD_0\Data\P0052318\
Data File : P0044998.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 May 2018 3:37
Operator : SM/UA
Sample : J3047-01
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
051818-SIDI-E-D-B

Manual Integrations
APPROVED

ugo
5/24/2018 9:10:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 06:13:13 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.016	3.288	4948898	5750965	32.295	23.990 #
2) SA Decachlor...	9.314	8.017	4519525	5967924	21.981	21.209m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\P0052318\
Data File : P0044998.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 May 2018 3:37
Operator : SM/UA
Sample : J3047-01
Misc :
ALS Vial : 40 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
051818-SIDI-E-D-B


Manual Integrations
APPROVED

ugo
5/24/2018 9:10:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 06:13:13 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.016 min
Delta R.T.: -0.017 min
Response: 4948898
Conc: 32.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
051818-SIDI-E-D-B


Manual Integrations
APPROVED

ugo
5/24/2018 9:10:22 PM




#1 Tetrachloro-m-xylene

R.T.: 3.288 min
Delta R.T.: -0.009 min
Response: 5750965
Conc: 23.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.314 min
Delta R.T.: -0.037 min
Response: 4519525
Conc: 21.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.017 min
Delta R.T.: -0.021 min
Response: 5967924
Conc: 21.21 ng/ml m