

Data Path : P:\HPCHEM1\ECD 0\Data\P0052418\
 Data File : P0045018.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 May 2018 10:49
 Operator : SM/UA
 Sample : J3106-01 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 A403696

Manual Integrations
 APPROVED

ugo
 5/29/2018 9:48:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 15:13:14 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 15 15:44:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.019	3.288	398792	574127	2.602m	2.395m
2) SA Decachlor...	9.318	8.020	557354	400294	2.711m	1.423m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\P0052418\
Data File : P0045018.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 May 2018 10:49
Operator : SM/UA
Sample : J3106-01 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
A403696


Manual Integrations
APPROVED

ugo
5/29/2018 9:48:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 15:13:14 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.019 min
Delta R.T.: -0.014 min
Response: 398792
Conc: 2.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
A403696


Manual Integrations
APPROVED

ugo
5/29/2018 9:48:38 PM




#1 Tetrachloro-m-xylene

R.T.: 3.288 min
Delta R.T.: -0.010 min
Response: 574127
Conc: 2.39 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.318 min
Delta R.T.: -0.032 min
Response: 557354
Conc: 2.71 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.020 min
Delta R.T.: -0.018 min
Response: 400294
Conc: 1.42 ng/ml m