

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0052518.M
 Title : GC EXTRACTABLES
 Last Update : Sat May 26 01:36:04 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0045067.D 1000 =P0045065.D 750 =P0045066.D
 250 =P0045068.D 50 =P0045069.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		6.083	5.533	5.836	6.223	5.403	5.816	E5	6.00
2) SA	Decachlorobiphenyl		5.568	5.114	5.267	5.735	4.639	5.264	E5	8.10
3) L1	AR-1016-1		1.233	1.113	1.154	1.314	1.218	1.207	E4	6.39
4) L1	AR-1016-2		1.538	1.411	1.446	1.645	1.522	1.512	E4	6.02
5) L1	AR-1016-3		1.265	1.115	1.181	1.334	1.249	1.229	E4	6.81
6) L1	AR-1016-4		1.274	1.051	1.190	1.321	1.149	1.197	E4	8.85
7) L1	AR-1016-5		1.080	0.961	1.021	1.120	1.141	1.065	E4	6.95
8) L2	AR-1221-1		1.173					1.173	E4	0.00
9) L2	AR-1221-2		8.767					8.767	E3	0.00
10) L2	AR-1221-3		2.790					2.790	E4	0.00
11) L3	AR-1232-1		1.268					1.268	E4	0.00
12) L3	AR-1232-2		8.830					8.830	E3	0.00
13) L3	AR-1232-3		1.467					1.467	E4	0.00
14) L3	AR-1232-4		8.638					8.638	E3	0.00
15) L3	AR-1232-5		6.942					6.942	E3	0.00
16) L4	AR-1242-1		1.163					1.163	E4	0.00
17) L4	AR-1242-2		2.497					2.497	E4	0.00
18) L4	AR-1242-3		1.203					1.203	E4	0.00
19) L4	AR-1242-4		1.124					1.124	E4	0.00
20) L4	AR-1242-5		1.280					1.280	E4	0.00
21) L5	AR-1248-1		1.889					1.889	E4	0.00
22) L5	AR-1248-2		1.667					1.667	E4	0.00
23) L5	AR-1248-3		2.210					2.210	E4	0.00
24) L5	AR-1248-4		2.148					2.148	E4	0.00
25) L5	AR-1248-5		1.475					1.475	E4	0.00
26) L6	AR-1254-1		3.442					3.442	E4	0.00
27) L6	AR-1254-2		2.244					2.244	E4	0.00
28) L6	AR-1254-3		3.843					3.843	E4	0.00
29) L6	AR-1254-4		3.192					3.192	E4	0.00
30) L6	AR-1254-5		2.304					2.304	E4	0.00
31) L7	AR-1260-1		2.408	2.208	2.219	2.498	2.316	2.330	E4	5.34
32) L7	AR-1260-2		3.082	2.820	2.903	3.206	3.084	3.019	E4	5.14
33) L7	AR-1260-3		3.827	3.539	3.650	4.116	3.730	3.772	E4	5.81
34) L7	AR-1260-4		6.634	6.306	6.337	6.595	5.804	6.335	E4	5.24
35) L7	AR-1260-5		3.976	3.746	3.736	4.016	3.550	3.805	E4	5.03
36) L8	AR-1262-1		2.379					2.379	E4	0.00
37) L8	AR-1262-2		3.127					3.127	E4	0.00
38) L8	AR-1262-3		3.013					3.013	E4	0.00
39) L8	AR-1262-4		4.133					4.133	E4	0.00
40) L8	AR-1262-5		9.388					9.388	E4	0.00
41) L9	AR-1268-1		1.004					1.004	E5	0.00
42) L9	AR-1268-2		8.907					8.907	E4	0.00
43) L9	AR-1268-3		7.730					7.730	E4	0.00
44) L9	AR-1268-4		3.204					3.204	E4	0.00
45) L9	AR-1268-5		2.350					2.350	E5	0.00

Signal #2 Calibration Files

500 =P0045067.D 1000 =P0045065.D 750 =P0045066.D
 250 =P0045068.D 50 =P0045069.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		6.681	5.945	6.314	6.971	6.104	6.403	E5	6.57
2) SA	Decachlorobiphenyl		6.115	5.774	5.837	6.182	4.756	5.733	E5	10.00

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0052518.M
 Title : GC EXTRACTABLES
 Last Update : Sat May 26 01:36:04 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0045067.D 1000 =P0045065.D 750 =P0045066.D
 250 =P0045068.D 50 =P0045069.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	1.391	1.218	1.295	1.476	1.334	1.343	E4	7.27
4)	L1	AR-1016-2	1.097	0.908	1.002	1.197	1.117	1.064	E4	10.48
5)	L1	AR-1016-3	1.343	1.120	1.224	1.447	1.259	1.279	E4	9.66
6)	L1	AR-1016-4	1.449	1.188	1.315	1.558	1.517	1.405	E4	10.86
7)	L1	AR-1016-5	1.398	1.165	1.257	1.503	1.368	1.338	E4	9.76
8)	L2	AR-1221-1	1.339					1.339	E4	0.00
9)	L2	AR-1221-2	9.543					9.543	E3	0.00
10)	L2	AR-1221-3	3.160					3.160	E4	0.00
11)	L3	AR-1232-1	1.461					1.461	E4	0.00
12)	L3	AR-1232-2	2.816					2.816	E4	0.00
13)	L3	AR-1232-3	1.090					1.090	E4	0.00
14)	L3	AR-1232-4	8.313					8.313	E3	0.00
15)	L3	AR-1232-5	7.180					7.180	E3	0.00
16)	L4	AR-1242-1	1.429					1.429	E4	0.00
17)	L4	AR-1242-2	4.929					4.929	E4	0.00
18)	L4	AR-1242-3	1.447					1.447	E4	0.00
19)	L4	AR-1242-4	1.354					1.354	E4	0.00
20)	L4	AR-1242-5	1.458					1.458	E4	0.00
21)	L5	AR-1248-1	2.452					2.452	E4	0.00
22)	L5	AR-1248-2	2.189					2.189	E4	0.00
23)	L5	AR-1248-3	3.555					3.555	E4	0.00
24)	L5	AR-1248-4	2.811					2.811	E4	0.00
25)	L5	AR-1248-5	1.554					1.554	E4	0.00
26)	L6	AR-1254-1	4.036					4.036	E4	0.00
27)	L6	AR-1254-2	3.474					3.474	E4	0.00
28)	L6	AR-1254-3	5.794					5.794	E4	0.00
29)	L6	AR-1254-4	4.452					4.452	E4	0.00
30)	L6	AR-1254-5	2.641					2.641	E4	0.00
31)	L7	AR-1260-1	3.143	2.735	2.866	3.457	3.143	3.069	E4	9.13
32)	L7	AR-1260-2	4.015	3.466	3.670	4.276	4.486	3.982	E4	10.54
33)	L7	AR-1260-3	4.305	3.852	4.030	4.530	4.383	4.220	E4	6.51
34)	L7	AR-1260-4	8.461	7.735	7.930	8.622	7.534	8.057	E4	5.81
35)	L7	AR-1260-5	5.844	5.383	5.520	5.967	5.197	5.582	E4	5.73
36)	L8	AR-1262-1	4.234					4.234	E4	0.00
37)	L8	AR-1262-2	3.778					3.778	E4	0.00
38)	L8	AR-1262-3	6.909					6.909	E4	0.00
39)	L8	AR-1262-4	5.671					5.671	E4	0.00
40)	L8	AR-1262-5	1.276					1.276	E5	0.00
41)	L9	AR-1268-1	1.344					1.344	E5	0.00
42)	L9	AR-1268-2	1.253					1.253	E5	0.00
43)	L9	AR-1268-3	1.079					1.079	E5	0.00
44)	L9	AR-1268-4	3.977					3.977	E4	0.00
45)	L9	AR-1268-5	2.989					2.989	E5	0.00

(#) = Out of Range