

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052720\
 Data File : P0068605.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 May 2020 16:57
 Operator : DD\AJ
 Sample : L2738-15
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 1723

Manual Integrations
 APPROVED

Sohil
 5/28/2020 4:09:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 17:57:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 08 13:29:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.904	3.948	517410	632696	18.448	18.438m
2) SA Decachlor...	10.837	9.203	633105	636175	15.910m	16.524

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052720\
Data File : P0068605.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 May 2020 16:57
Operator : DD\AJ
Sample : L2738-15
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
1723


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 17:57:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 08 13:29:29 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.904 min
Delta R.T.: -0.005 min
Response: 517410
Conc: 18.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
1723


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:24 PM




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: 0.015 min
Response: 632696
Conc: 18.44 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.837 min
Delta R.T.: -0.036 min
Response: 633105
Conc: 15.91 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.203 min
Delta R.T.: -0.007 min
Response: 636175
Conc: 16.52 ng/ml