

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052720\
 Data File : P0068619.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 May 2020 21:14
 Operator : DD\AJ
 Sample : L2746-19
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 NM71-COMP-2020-0-6

Manual Integrations
 APPROVED

Sohil
 5/28/2020 4:09:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 03:37:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0050820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 08 13:29:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.894	3.931	617441	796853	22.015	23.222
2) SA Decachlor...	10.837	9.200	469744	545235	11.805m	14.162m
Target Compounds						
31) L7 AR-1260-1	7.891	6.765	228672	376683	164.130m	196.264
32) L7 AR-1260-2	8.158	6.963	699929	464534	394.492	198.001 #
33) L7 AR-1260-3	8.518	7.115	282312	336915	194.329	154.603
34) L7 AR-1260-4	8.748	7.596	290672	320099	186.667	168.269
35) L7 AR-1260-5	9.068	7.844	565327	748186	170.189	168.351

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052720\
Data File : P0068619.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 May 2020 21:14
Operator : DD\AJ
Sample : L2746-19
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
NM71-COMP-2020-0-6


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 03:37:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 08 13:29:29 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.894 min
Delta R.T.: -0.015 min
Response: 617441
Conc: 22.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
NM71-COMP-2020-0-6


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:34 PM




#1 Tetrachloro-m-xylene

R.T.: 3.931 min
Delta R.T.: -0.002 min
Response: 796853
Conc: 23.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.837 min
Delta R.T.: -0.037 min
Response: 469744
Conc: 11.80 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.200 min
Delta R.T.: -0.010 min
Response: 545235
Conc: 14.16 ng/ml m




#31 AR-1260-1

R.T.: 7.891 min
Delta R.T.: -0.020 min
Response: 228672
Conc: 164.13 ng/ml m

Instrument :
ECD_O
ClientSampleId :
NM71-COMP-2020-0-6


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:34 PM




#31 AR-1260-1

R.T.: 6.765 min
Delta R.T.: -0.007 min
Response: 376683
Conc: 196.26 ng/ml




#32 AR-1260-2

R.T.: 8.158 min
Delta R.T.: -0.018 min
Response: 699929
Conc: 394.49 ng/ml



#32 AR-1260-2

R.T.: 6.963 min
Delta R.T.: -0.006 min
Response: 464534
Conc: 198.00 ng/ml




#33 AR-1260-3

R.T.: 8.518 min
Delta R.T.: -0.021 min
Response: 282312
Conc: 194.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
NM71-COMP-2020-0-6


Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:34 PM




#33 AR-1260-3

R.T.: 7.115 min
Delta R.T.: -0.007 min
Response: 336915
Conc: 154.60 ng/ml




#34 AR-1260-4

R.T.: 8.748 min
Delta R.T.: -0.020 min
Response: 290672
Conc: 186.67 ng/ml



#34 AR-1260-4

R.T.: 7.596 min
Delta R.T.: -0.008 min
Response: 320099
Conc: 168.27 ng/ml




#35 AR-1260-5

R.T.: 9.068 min
Delta R.T.: -0.021 min
Response: 565327
Conc: 170.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
NM71-COMP-2020-0-6

Manual Integrations
APPROVED

Sohil
5/28/2020 4:09:34 PM



#35 AR-1260-5

R.T.: 7.844 min
Delta R.T.: -0.007 min
Response: 748186
Conc: 168.35 ng/ml