

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052820\
 Data File : P0068637.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2020 15:50
 Operator : DD\AJ
 Sample : P0052820ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 16:12:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0052820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 28 16:02:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	31.218	31.594 E3	-1.2	98	0.00
2 SA Decachlorobiphenyl	43.854	42.800 E3	2.4	98	0.00
3 L1 AR-1016-1	1.328	1.242 E3	6.5	97	0.00
4 L1 AR-1016-2	1.845	1.751 E3	5.1	98	0.00
5 L1 AR-1016-3	1.148	1.042 E3	9.2	97	0.00
6 L1 AR-1016-4	908.069	905.370	0.3	102	0.00
7 L1 AR-1016-5	839.385	871.552	-3.8	102	0.00
31 L7 AR-1260-1	1.720	1.714 E3	0.3	101	0.00
32 L7 AR-1260-2	2.303	2.340 E3	-1.6	101	0.00
33 L7 AR-1260-3	1.937	1.948 E3	-0.6	100	0.00
34 L7 AR-1260-4	1.854	1.895 E3	-2.2	101	0.00
35 L7 AR-1260-5	4.240	4.347 E3	-2.5	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	36.978	37.272 E3	-0.8	102	0.00
2 SA Decachlorobiphenyl	42.604	44.250 E3	-3.9	102	0.00
3 L1 AR-1016-1	1.528	1.558 E3	-2.0	103	0.00
4 L1 AR-1016-2	2.056	2.081 E3	-1.2	102	0.00
5 L1 AR-1016-3	1.106	1.123 E3	-1.5	103	0.00
6 L1 AR-1016-4	893.772	911.952	-2.0	102	0.00
7 L1 AR-1016-5	1.140	1.180 E3	-3.5	103	0.00
31 L7 AR-1260-1	2.149	2.149 E3	0.0	103	0.00
32 L7 AR-1260-2	2.679	2.698 E3	-0.7	104	0.00
33 L7 AR-1260-3	2.422	2.436 E3	-0.6	104	0.00
34 L7 AR-1260-4	2.047	2.111 E3	-3.1	105	0.00
35 L7 AR-1260-5	5.068	5.149 E3	-1.6	102	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	317.270	0.000	100.0#	0#	-5.14#
9 L2 AR-1221-2	227.355	0.000	100.0#	0#	-5.24#
10 L2 AR-1221-3	804.955	0.000	100.0#	0#	-5.33#
11 L3 AR-1232-1	652.742	0.000	100.0#	0#	-5.33#
12 L3 AR-1232-2	333.332	0.000	100.0#	0#	-5.92#
13 L3 AR-1232-3	749.363	0.000	100.0#	0#	-6.24#
14 L3 AR-1232-4	358.442	0.000	100.0#	0#	-6.41#
15 L3 AR-1232-5	268.429	0.000	100.0#	0#	-6.51#
16 L4 AR-1242-1	988.081	0.000	100.0#	0#	-6.21#
17 L4 AR-1242-2	1.390	0.000 E3	100.0#	0#	-6.24#
18 L4 AR-1242-3	826.594	0.000	100.0#	0#	-6.30#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052820\
 Data File : P0068637.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2020 15:50
 Operator : DD\AJ
 Sample : P0052820ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 16:12:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0052820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 28 16:02:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	703.488	0.000	100.0#	0#	-6.41#
20 L4 AR-1242-5	859.171	0.000	100.0#	0#	-7.19#
21 L5 AR-1248-1	887.131	0.000	100.0#	0#	-6.24#
22 L5 AR-1248-2	1.067	0.000 E3	100.0#	0#	-6.51#
23 L5 AR-1248-3	1.178	0.000 E3	100.0#	0#	-6.72#
24 L5 AR-1248-4	1.553	0.000 E3	100.0#	0#	-7.15#
25 L5 AR-1248-5	1.482	0.000 E3	100.0#	0#	-7.19#
26 L6 AR-1254-1	1.513	0.000 E3	100.0#	0#	-7.12#
27 L6 AR-1254-2	2.381	0.000 E3	100.0#	0#	-7.35#
28 L6 AR-1254-3	2.486	0.000 E3	100.0#	0#	-7.73#
29 L6 AR-1254-4	2.119	0.000 E3	100.0#	0#	-8.02#
30 L6 AR-1254-5	2.084	0.000 E3	100.0#	0#	-8.45#
36 L8 AR-1262-1	3.119	0.000 E3	100.0#	0#	-8.52#
37 L8 AR-1262-2	5.182	0.000 E3	100.0#	0#	-9.07#
38 L8 AR-1262-3	3.394	0.000 E3	100.0#	0#	-9.37#
39 L8 AR-1262-4	2.627	0.000 E3	100.0#	0#	-9.46#
40 L8 AR-1262-5	1.947	0.000 E3	100.0#	0#	-10.11#
41 L9 AR-1268-1	6.633	0.000 E3	100.0#	0#	-9.37#
42 L9 AR-1268-2	5.900	0.000 E3	100.0#	0#	-9.46#
43 L9 AR-1268-3	5.092	0.000 E3	100.0#	0#	-9.68#
44 L9 AR-1268-4	2.299	0.000 E3	100.0#	0#	-10.11#
45 L9 AR-1268-5	16.050	0.000 E3	100.0#	0#	-10.51#

Signal #2

8 L2 AR-1221-1	432.233	0.000	100.0#	0#	-4.19#
9 L2 AR-1221-2	327.624	0.000	100.0#	0#	-4.28#
10 L2 AR-1221-3	1.004	0.000 E3	100.0#	0#	-4.37#
11 L3 AR-1232-1	820.750	0.000	100.0#	0#	-4.37#
12 L3 AR-1232-2	891.541	0.000	100.0#	0#	-5.20#
13 L3 AR-1232-3	479.411	0.000	100.0#	0#	-5.39#
14 L3 AR-1232-4	410.659	0.000	100.0#	0#	-5.49#
15 L3 AR-1232-5	460.815	0.000	100.0#	0#	-5.67#
16 L4 AR-1242-1	1.227	0.000 E3	100.0#	0#	-5.18#
17 L4 AR-1242-2	1.640	0.000 E3	100.0#	0#	-5.20#
18 L4 AR-1242-3	886.267	0.000	100.0#	0#	-5.39#
19 L4 AR-1242-4	843.996	0.000	100.0#	0#	-5.49#
20 L4 AR-1242-5	1.157	0.000 E3	100.0#	0#	-6.05#
21 L5 AR-1248-1	965.986	0.000	100.0#	0#	-5.18#
22 L5 AR-1248-2	1.247	0.000 E3	100.0#	0#	-5.44#
23 L5 AR-1248-3	1.235	0.000 E3	100.0#	0#	-5.49#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052820\
 Data File : P0068637.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2020 15:50
 Operator : DD\AJ
 Sample : P0052820ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 16:12:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0052820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 28 16:02:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.539	0.000 E3	100.0#	0#	-5.67#
25 L5 AR-1248-5	1.569	0.000 E3	100.0#	0#	-6.09#
26 L6 AR-1254-1	2.437	0.000 E3	100.0#	0#	-6.05#
27 L6 AR-1254-2	2.150	0.000 E3	100.0#	0#	-6.21#
28 L6 AR-1254-3	3.412	0.000 E3	100.0#	0#	-6.63#
29 L6 AR-1254-4	2.347	0.000 E3	100.0#	0#	-6.87#
30 L6 AR-1254-5	3.045	0.000 E3	100.0#	0#	-7.29#
36 L8 AR-1262-1	1.673	0.000 E3	100.0#	0#	-7.29#
37 L8 AR-1262-2	5.684	0.000 E3	100.0#	0#	-7.84#
38 L8 AR-1262-3	2.341	0.000 E3	100.0#	0#	-8.13#
39 L8 AR-1262-4	4.238	0.000 E3	100.0#	0#	-8.19#
40 L8 AR-1262-5	2.107	0.000 E3	100.0#	0#	-8.69#
41 L9 AR-1268-1	6.818	0.000 E3	100.0#	0#	-8.13#
42 L9 AR-1268-2	6.276	0.000 E3	100.0#	0#	-8.19#
43 L9 AR-1268-3	5.204	0.000 E3	100.0#	0#	-8.40#
44 L9 AR-1268-4	2.408	0.000 E3	100.0#	0#	-8.69#
45 L9 AR-1268-5	16.268	0.000 E3	100.0#	0#	-8.96#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0