

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052820\
 Data File : P0068637.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2020 15:50
 Operator : DD\AJ
 Sample : P0052820ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO052820

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 29 02:31:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 29 02:25:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.892	3.929	1579683	1863621	50.959	50.446
2)	SA Decachlor...	10.838	9.200	2139976	2203785	49.625	51.899
Target Compounds							
3)	L1 AR-1016-1	6.213	5.181	636634	772483	507.129	504.547
4)	L1 AR-1016-2	6.237	5.201	891824	1036228	501.436	503.346
5)	L1 AR-1016-3	6.302	5.390	536446	561414	509.681	520.318
6)	L1 AR-1016-4	6.409	5.443	452685	455976	519.144	518.258
7)	L1 AR-1016-5	6.722	5.670	435746	590053	519.125	517.560
31)	L7 AR-1260-1	7.892	6.764	858368	1074733	499.004	500.142
32)	L7 AR-1260-2	8.157	6.962	1169843	1348994	508.050	503.534
33)	L7 AR-1260-3	8.519	7.115	973851	1218133	502.790	503.037
34)	L7 AR-1260-4	8.749	7.596	945059	1055283	509.642	515.452
35)	L7 AR-1260-5	9.069	7.844	2172482	2574345	512.367	507.981



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052820\
Data File : P0068637.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 May 2020 15:50
Operator : DD\AJ
Sample : P0052820ICV500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO052820

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 29 02:31:00 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 29 02:25:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.892 min
Delta R.T.: 0.000 min
Response: 1579683
Conc: 50.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#1 Tetrachloro-m-xylene

R.T.: 3.929 min
Delta R.T.: 0.000 min
Response: 1863621
Conc: 50.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.838 min
Delta R.T.: 0.000 min
Response: 2139976
Conc: 49.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.200 min
Delta R.T.: 0.000 min
Response: 2203785
Conc: 51.90 ng/ml



#3 AR-1016-1


R.T.: 6.213 min
Delta R.T.: 0.000 min
Response: 636634
Conc: 507.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#3 AR-1016-1

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 772483
Conc: 504.55 ng/ml




#4 AR-1016-2

R.T.: 6.237 min
Delta R.T.: 0.000 min
Response: 891824
Conc: 501.44 ng/ml



#4 AR-1016-2


R.T.: 5.201 min
Delta R.T.: 0.000 min
Response: 1036228
Conc: 503.35 ng/ml



#5 AR-1016-3


R.T.: 6.302 min
Delta R.T.: 0.000 min
Response: 536446
Conc: 509.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#5 AR-1016-3

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 561414
Conc: 520.32 ng/ml




#6 AR-1016-4

R.T.: 6.409 min
Delta R.T.: 0.000 min
Response: 452685
Conc: 519.14 ng/ml



#6 AR-1016-4


R.T.: 5.443 min
Delta R.T.: 0.000 min
Response: 455976
Conc: 518.26 ng/ml



#7 AR-1016-5


R.T.: 6.722 min
Delta R.T.: 0.000 min
Response: 435746
Conc: 519.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#7 AR-1016-5

R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 590053
Conc: 517.56 ng/ml




#31 AR-1260-1

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 858368
Conc: 499.00 ng/ml



#31 AR-1260-1


R.T.: 6.764 min
Delta R.T.: 0.000 min
Response: 1074733
Conc: 500.14 ng/ml



#32 AR-1260-2


R.T.: 8.157 min
Delta R.T.: 0.002 min
Response: 1169843
Conc: 508.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#32 AR-1260-2

R.T.: 6.962 min
Delta R.T.: 0.000 min
Response: 1348994
Conc: 503.53 ng/ml




#33 AR-1260-3

R.T.: 8.519 min
Delta R.T.: 0.000 min
Response: 973851
Conc: 502.79 ng/ml



#33 AR-1260-3


R.T.: 7.115 min
Delta R.T.: 0.000 min
Response: 1218133
Conc: 503.04 ng/ml



#34 AR-1260-4


R.T.: 8.749 min
Delta R.T.: 0.000 min
Response: 945059
Conc: 509.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO052820




#34 AR-1260-4

R.T.: 7.596 min
Delta R.T.: 0.000 min
Response: 1055283
Conc: 515.45 ng/ml



#35 AR-1260-5

R.T.: 9.069 min
Delta R.T.: 0.001 min
Response: 2172482
Conc: 512.37 ng/ml



#35 AR-1260-5

R.T.: 7.844 min
Delta R.T.: 0.000 min
Response: 2574345
Conc: 507.98 ng/ml