

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052820\
 Data File : P0068648.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2020 19:07
 Operator : DD\AJ
 Sample : L2791-06
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SIMW-04-20200527

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 29 02:34:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 29 02:25:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.893	3.931	777678	846731	25.087	22.920
2) SA Decachlor...	10.839	9.201	819872	876651	19.012	20.645

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052820\
Data File : P0068648.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 May 2020 19:07
Operator : DD\AJ
Sample : L2791-06
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SIMW-04-20200527

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 29 02:34:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 29 02:25:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.893 min
Delta R.T.: 0.001 min
Response: 777678
Conc: 25.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
SIMW-04-20200527




#1 Tetrachloro-m-xylene

R.T.: 3.931 min
Delta R.T.: 0.002 min
Response: 846731
Conc: 22.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.839 min
Delta R.T.: 0.001 min
Response: 819872
Conc: 19.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.201 min
Delta R.T.: 0.000 min
Response: 876651
Conc: 20.65 ng/ml