

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0052820\  
 Data File : P0068672.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 29 May 2020 2:49  
 Operator : DD\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: May 29 04:16:34 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0052820.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri May 29 02:25:52 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1 | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|--------|-------|--------|---------|--------|--------|
| -----                       |        |       |        |         |        |        |
| System Monitoring Compounds |        |       |        |         |        |        |
| 1) SA Tetrachlo...          | 4.890  | 3.929 | 748238 | 936940  | 24.137 | 25.362 |
| 2) SA Decachlor...          | 10.835 | 9.199 | 945004 | 1073734 | 21.914 | 25.287 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0052820\  
Data File : P0068672.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 29 May 2020 2:49  
Operator : DD\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: May 29 04:16:34 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0052820.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri May 29 02:25:52 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.890 min  
Delta R.T.: -0.002 min  
Response: 748238  
Conc: 24.14 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.929 min  
Delta R.T.: 0.000 min  
Response: 936940  
Conc: 25.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.835 min  
Delta R.T.: -0.003 min  
Response: 945004  
Conc: 21.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.199 min  
Delta R.T.: -0.002 min  
Response: 1073734  
Conc: 25.29 ng/ml