

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052919\
 Data File : P0056702.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 May 2019 21:26
 Operator : SM/SJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 5/30/2019 9:31:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 01:19:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.251	3.569	2108798	1700223	55.084	51.556
2)	SA Decachlor...	9.828	8.504	2639671	1914967	53.779	53.706
Target Compounds							
3)	L1 AR-1016-1	5.413	4.638	851003	615492	473.142m	489.074
4)	L1 AR-1016-2	5.434	4.656	1276952	889063	508.528m	493.479
5)	L1 AR-1016-3	5.496	4.830	818352	493700	509.148	500.738
6)	L1 AR-1016-4	5.594	4.871	675570	416642	514.995	503.287
7)	L1 AR-1016-5	5.885	5.081	687515	540328	497.619	503.698m
31)	L7 AR-1260-1	7.001	6.103	1254488	1018008	488.855	508.907
32)	L7 AR-1260-2	7.258	6.291	1578907	1297801	522.097	526.500
33)	L7 AR-1260-3	7.614	6.442	1279021	1180177	530.923	526.107
34)	L7 AR-1260-4	7.839	6.910	1414812	1026804	540.167	551.742
35)	L7 AR-1260-5	8.148	7.151	2670493	2489843	558.437	577.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052919\
Data File : PO056702.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 May 2019 21:26
Operator : SM/SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 01:19:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.251 min
Delta R.T.: -0.005 min
Response: 2108798
Conc: 55.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#1 Tetrachloro-m-xylene

R.T.: 3.569 min
Delta R.T.: -0.004 min
Response: 1700223
Conc: 51.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.828 min
Delta R.T.: -0.008 min
Response: 2639671
Conc: 53.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.504 min
Delta R.T.: -0.009 min
Response: 1914967
Conc: 53.71 ng/ml




#3 AR-1016-1

R.T.: 5.413 min
Delta R.T.: -0.005 min
Response: 851003
Conc: 473.14 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#3 AR-1016-1

R.T.: 4.638 min
Delta R.T.: -0.005 min
Response: 615492
Conc: 489.07 ng/ml




#4 AR-1016-2

R.T.: 5.434 min
Delta R.T.: -0.006 min
Response: 1276952
Conc: 508.53 ng/ml m



#4 AR-1016-2

R.T.: 4.656 min
Delta R.T.: -0.006 min
Response: 889063
Conc: 493.48 ng/ml




#5 AR-1016-3

R.T.: 5.496 min
Delta R.T.: -0.004 min
Response: 818352
Conc: 509.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#5 AR-1016-3

R.T.: 4.830 min
Delta R.T.: -0.006 min
Response: 493700
Conc: 500.74 ng/ml




#6 AR-1016-4

R.T.: 5.594 min
Delta R.T.: -0.005 min
Response: 675570
Conc: 515.00 ng/ml



#6 AR-1016-4

R.T.: 4.871 min
Delta R.T.: -0.005 min
Response: 416642
Conc: 503.29 ng/ml




#7 AR-1016-5

R.T.: 5.885 min
Delta R.T.: -0.005 min
Response: 687515
Conc: 497.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#7 AR-1016-5

R.T.: 5.081 min
Delta R.T.: -0.007 min
Response: 540328
Conc: 503.70 ng/ml m




#31 AR-1260-1

R.T.: 7.001 min
Delta R.T.: -0.006 min
Response: 1254488
Conc: 488.85 ng/ml



#31 AR-1260-1

R.T.: 6.103 min
Delta R.T.: -0.007 min
Response: 1018008
Conc: 508.91 ng/ml




#32 AR-1260-2

R.T.: 7.258 min
Delta R.T.: -0.006 min
Response: 1578907
Conc: 522.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#32 AR-1260-2

R.T.: 6.291 min
Delta R.T.: -0.007 min
Response: 1297801
Conc: 526.50 ng/ml




#33 AR-1260-3

R.T.: 7.614 min
Delta R.T.: -0.006 min
Response: 1279021
Conc: 530.92 ng/ml



#33 AR-1260-3

R.T.: 6.442 min
Delta R.T.: -0.007 min
Response: 1180177
Conc: 526.11 ng/ml




#34 AR-1260-4

R.T.: 7.839 min
Delta R.T.: -0.006 min
Response: 1414812
Conc: 540.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/30/2019 9:31:57 AM




#34 AR-1260-4

R.T.: 6.910 min
Delta R.T.: -0.007 min
Response: 1026804
Conc: 551.74 ng/ml



#35 AR-1260-5

R.T.: 8.148 min
Delta R.T.: -0.006 min
Response: 2670493
Conc: 558.44 ng/ml



#35 AR-1260-5

R.T.: 7.151 min
Delta R.T.: -0.007 min
Response: 2489843
Conc: 577.38 ng/ml