

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060222\
 Data File : P0086854.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jun 2022 10:47
 Operator : YP\AJ
 Sample : PB145238BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 02:07:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 17:55:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.515	3.666	1299751	639882	22.726	20.018
2)	SA Decachlor...	10.391	8.697	1095064	519565	23.991	23.171
Target Compounds							
3)	L1 AR-1016-1	5.699	4.760	1061600	456090	534.636	462.045
4)	L1 AR-1016-2	5.721	4.779	1514564	628698	534.483	460.980
5)	L1 AR-1016-3	5.784	4.955	948814	340011	542.719	461.243
6)	L1 AR-1016-4	5.884	4.998	759197	265914	543.542	455.265
7)	L1 AR-1016-5	6.181	5.212	747654	337552	544.124	463.598
8)	L2 AR-1221-1	4.723	3.882	127596	54177	192.066	141.929 #
9)	L2 AR-1221-2	4.813	3.969	125030	68476	238.588	241.695
10)	L2 AR-1221-3	4.891	4.046	576026	277360	357.073	329.307
11)	L3 AR-1232-1	4.891	4.046	576026	277360	377.223	342.707
12)	L3 AR-1232-2	5.232	4.779	815220	628698	1159.912	949.026
13)	L3 AR-1232-3	5.721	4.955	1514564	340011	1116.997	986.431
14)	L3 AR-1232-4	5.884	5.040	759197	328520	1162.325	1083.827
15)	L3 AR-1232-5	5.975	5.212	683676	337552	1304.843	1021.551
16)	L4 AR-1242-1	5.699	4.760	1061600	456090	664.122	580.574
17)	L4 AR-1242-2	5.721	4.779	1514564	628698	664.807	581.652
18)	L4 AR-1242-3	5.784	4.955	948814	340011	668.056	575.899
19)	L4 AR-1242-4	5.884	5.040	759197	328520	669.870	574.110
20)	L4 AR-1242-5	6.628	5.565	100514	264633	85.476	394.041 #
21)	L5 AR-1248-1	5.699	4.760	1061600	456090	870.547	775.701
22)	L5 AR-1248-2	5.975	4.998	683676	265914	390.850	330.313
23)	L5 AR-1248-3	6.181	5.040	747654	328520	386.847	391.532
24)	L5 AR-1248-4	6.587	5.212	99491	337552	46.322	346.979 #
25)	L5 AR-1248-5	6.628	5.604	100514	40613	49.037	43.689
26)	L6 AR-1254-1	6.563	5.565	581258	264633	261.599	184.007 #
27)	L6 AR-1254-2	6.782	5.713	619727	253777	180.884	204.833
28)	L6 AR-1254-3	7.152	6.133	308950	462122	88.431	230.055 #
29)	L6 AR-1254-4	7.439	6.345	202699	50588	79.022	40.994 #
30)	L6 AR-1254-5	7.860	6.763	1934570	940425	699.428	542.933
31)	L7 AR-1260-1	7.317	6.248	1363859	651131	549.544	502.810
32)	L7 AR-1260-2	7.575	6.435	1617176	781169	477.016	505.747
33)	L7 AR-1260-3	7.937	6.589	1066193	757532	493.432	435.897
34)	L7 AR-1260-4	8.167	7.062	1292077	552854	502.822	485.730
35)	L7 AR-1260-5	8.496	7.303	2358636	1311956	496.357	493.995
36)	L8 AR-1262-1	7.937	6.803	1066193	570855	321.017	330.134
37)	L8 AR-1262-2	8.496	7.303	2358636	1311956	412.230	436.877
38)	L8 AR-1262-3	8.834	7.589	1546329	248150	392.710	206.982 #
39)	L8 AR-1262-4	8.913	7.650	306307	947018	102.278	433.283 #
40)	L8 AR-1262-5	9.598	8.146	592046	282777	286.322	280.366
41)	L9 AR-1268-1	8.834	7.589	1546329	248150	220.148	72.413 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060222\
 Data File : P0086854.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jun 2022 10:47
 Operator : YP\AJ
 Sample : PB145238BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 02:07:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 17:55:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.913	7.650	306307	947018	47.840	301.976 #
43) L9	AR-1268-3	9.157	7.861	47069	24594	8.654	9.293
44) L9	AR-1268-4	9.598	8.146	592046	282777	250.711	247.546
45) L9	AR-1268-5	10.032	8.440	195631	92056	10.873	10.540



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060222\
Data File : P0086854.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jun 2022 10:47
Operator : YP\AJ
Sample : PB145238BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 02:07:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 17:55:39 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 1299751
Conc: 22.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 639882
Conc: 20.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.391 min
Delta R.T.: 0.008 min
Response: 1095064
Conc: 23.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.697 min
Delta R.T.: -0.003 min
Response: 519565
Conc: 23.17 ng/ml



#3 AR-1016-1


R.T.: 5.699 min
Delta R.T.: 0.001 min
Response: 1061600
Conc: 534.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 456090
Conc: 462.05 ng/ml




#4 AR-1016-2

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 1514564
Conc: 534.48 ng/ml



#4 AR-1016-2


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 628698
Conc: 460.98 ng/ml



#5 AR-1016-3


R.T.: 5.784 min
Delta R.T.: 0.002 min
Response: 948814
Conc: 542.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.955 min
Delta R.T.: -0.001 min
Response: 340011
Conc: 461.24 ng/ml




#6 AR-1016-4

R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 759197
Conc: 543.54 ng/ml



#6 AR-1016-4


R.T.: 4.998 min
Delta R.T.: -0.002 min
Response: 265914
Conc: 455.26 ng/ml



#7 AR-1016-5


R.T.: 6.181 min
Delta R.T.: 0.000 min
Response: 747654
Conc: 544.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.212 min
Delta R.T.: -0.002 min
Response: 337552
Conc: 463.60 ng/ml




#8 AR-1221-1

R.T.: 4.723 min
Delta R.T.: -0.002 min
Response: 127596
Conc: 192.07 ng/ml



#8 AR-1221-1


R.T.: 3.882 min
Delta R.T.: -0.003 min
Response: 54177
Conc: 141.93 ng/ml



#9 AR-1221-2


R.T.: 4.813 min
Delta R.T.: 0.000 min
Response: 125030
Conc: 238.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.969 min
Delta R.T.: -0.002 min
Response: 68476
Conc: 241.69 ng/ml




#10 AR-1221-3

R.T.: 4.891 min
Delta R.T.: 0.000 min
Response: 576026
Conc: 357.07 ng/ml



#10 AR-1221-3


R.T.: 4.046 min
Delta R.T.: -0.002 min
Response: 277360
Conc: 329.31 ng/ml



#11 AR-1232-1


R.T.: 4.891 min
Delta R.T.: 0.000 min
Response: 576026
Conc: 377.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.046 min
Delta R.T.: -0.002 min
Response: 277360
Conc: 342.71 ng/ml




#12 AR-1232-2

R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 815220
Conc: 1159.91 ng/ml



#12 AR-1232-2


R.T.: 4.779 min
Delta R.T.: -0.002 min
Response: 628698
Conc: 949.03 ng/ml



#13 AR-1232-3


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 1514564
Conc: 1117.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.955 min
Delta R.T.: -0.002 min
Response: 340011
Conc: 986.43 ng/ml




#14 AR-1232-4

R.T.: 5.884 min
Delta R.T.: 0.001 min
Response: 759197
Conc: 1162.32 ng/ml



#14 AR-1232-4


R.T.: 5.040 min
Delta R.T.: -0.002 min
Response: 328520
Conc: 1083.83 ng/ml



#15 AR-1232-5


R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 683676
Conc: 1304.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.212 min
Delta R.T.: -0.002 min
Response: 337552
Conc: 1021.55 ng/ml




#16 AR-1242-1

R.T.: 5.699 min
Delta R.T.: 0.000 min
Response: 1061600
Conc: 664.12 ng/ml



#16 AR-1242-1


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 456090
Conc: 580.57 ng/ml



#17 AR-1242-2


R.T.: 5.721 min
Delta R.T.: -0.001 min
Response: 1514564
Conc: 664.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.779 min
Delta R.T.: -0.002 min
Response: 628698
Conc: 581.65 ng/ml




#18 AR-1242-3

R.T.: 5.784 min
Delta R.T.: -0.002 min
Response: 948814
Conc: 668.06 ng/ml



#18 AR-1242-3


R.T.: 4.955 min
Delta R.T.: -0.001 min
Response: 340011
Conc: 575.90 ng/ml



#19 AR-1242-4


R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 759197
Conc: 669.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.040 min
Delta R.T.: -0.001 min
Response: 328520
Conc: 574.11 ng/ml




#20 AR-1242-5

R.T.: 6.628 min
Delta R.T.: 0.000 min
Response: 100514
Conc: 85.48 ng/ml



#20 AR-1242-5


R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 264633
Conc: 394.04 ng/ml



#21 AR-1248-1


R.T.: 5.699 min
Delta R.T.: 0.000 min
Response: 1061600
Conc: 870.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.760 min
Delta R.T.: -0.001 min
Response: 456090
Conc: 775.70 ng/ml




#22 AR-1248-2

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 683676
Conc: 390.85 ng/ml



#22 AR-1248-2


R.T.: 4.998 min
Delta R.T.: -0.002 min
Response: 265914
Conc: 330.31 ng/ml



#23 AR-1248-3


R.T.: 6.181 min
Delta R.T.: 0.000 min
Response: 747654
Conc: 386.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.040 min
Delta R.T.: -0.001 min
Response: 328520
Conc: 391.53 ng/ml




#24 AR-1248-4

R.T.: 6.587 min
Delta R.T.: -0.002 min
Response: 99491
Conc: 46.32 ng/ml



#24 AR-1248-4


R.T.: 5.212 min
Delta R.T.: -0.001 min
Response: 337552
Conc: 346.98 ng/ml



#25 AR-1248-5


R.T.: 6.628 min
Delta R.T.: 0.000 min
Response: 100514
Conc: 49.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.604 min
Delta R.T.: -0.002 min
Response: 40613
Conc: 43.69 ng/ml




#26 AR-1254-1

R.T.: 6.563 min
Delta R.T.: -0.001 min
Response: 581258
Conc: 261.60 ng/ml



#26 AR-1254-1


R.T.: 5.565 min
Delta R.T.: -0.002 min
Response: 264633
Conc: 184.01 ng/ml



#27 AR-1254-2


R.T.: 6.782 min
Delta R.T.: -0.003 min
Response: 619727
Conc: 180.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.713 min
Delta R.T.: -0.001 min
Response: 253777
Conc: 204.83 ng/ml




#28 AR-1254-3

R.T.: 7.152 min
Delta R.T.: -0.002 min
Response: 308950
Conc: 88.43 ng/ml



#28 AR-1254-3


R.T.: 6.133 min
Delta R.T.: 0.016 min
Response: 462122
Conc: 230.06 ng/ml



#29 AR-1254-4


R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 202699
Conc: 79.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.345 min
Delta R.T.: -0.002 min
Response: 50588
Conc: 40.99 ng/ml




#30 AR-1254-5

R.T.: 7.860 min
Delta R.T.: -0.003 min
Response: 1934570
Conc: 699.43 ng/ml



#30 AR-1254-5


R.T.: 6.763 min
Delta R.T.: -0.003 min
Response: 940425
Conc: 542.93 ng/ml



#31 AR-1260-1


R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 1363859
Conc: 549.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.248 min
Delta R.T.: 0.000 min
Response: 651131
Conc: 502.81 ng/ml




#32 AR-1260-2

R.T.: 7.575 min
Delta R.T.: -0.003 min
Response: 1617176
Conc: 477.02 ng/ml



#32 AR-1260-2


R.T.: 6.435 min
Delta R.T.: -0.002 min
Response: 781169
Conc: 505.75 ng/ml



#33 AR-1260-3


R.T.: 7.937 min
Delta R.T.: 0.000 min
Response: 1066193
Conc: 493.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.589 min
Delta R.T.: 0.000 min
Response: 757532
Conc: 435.90 ng/ml




#34 AR-1260-4

R.T.: 8.167 min
Delta R.T.: 0.000 min
Response: 1292077
Conc: 502.82 ng/ml



#34 AR-1260-4


R.T.: 7.062 min
Delta R.T.: -0.001 min
Response: 552854
Conc: 485.73 ng/ml



#35 AR-1260-5


R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 2358636
Conc: 496.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.002 min
Response: 1311956
Conc: 493.99 ng/ml




#36 AR-1262-1

R.T.: 7.937 min
Delta R.T.: -0.001 min
Response: 1066193
Conc: 321.02 ng/ml



#36 AR-1262-1


R.T.: 6.803 min
Delta R.T.: -0.002 min
Response: 570855
Conc: 330.13 ng/ml



#37 AR-1262-2


R.T.: 8.496 min
Delta R.T.: -0.001 min
Response: 2358636
Conc: 412.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.303 min
Delta R.T.: -0.002 min
Response: 1311956
Conc: 436.88 ng/ml




#38 AR-1262-3

R.T.: 8.834 min
Delta R.T.: 0.011 min
Response: 1546329
Conc: 392.71 ng/ml



#38 AR-1262-3


R.T.: 7.589 min
Delta R.T.: -0.001 min
Response: 248150
Conc: 206.98 ng/ml



#39 AR-1262-4


R.T.: 8.913 min
Delta R.T.: -0.002 min
Response: 306307
Conc: 102.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.650 min
Delta R.T.: -0.003 min
Response: 947018
Conc: 433.28 ng/ml




#40 AR-1262-5

R.T.: 9.598 min
Delta R.T.: 0.000 min
Response: 592046
Conc: 286.32 ng/ml



#40 AR-1262-5


R.T.: 8.146 min
Delta R.T.: -0.002 min
Response: 282777
Conc: 280.37 ng/ml



#41 AR-1268-1


R.T.: 8.834 min
Delta R.T.: 0.012 min
Response: 1546329
Conc: 220.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.589 min
Delta R.T.: -0.002 min
Response: 248150
Conc: 72.41 ng/ml




#42 AR-1268-2

R.T.: 8.913 min
Delta R.T.: -0.007 min
Response: 306307
Conc: 47.84 ng/ml



#42 AR-1268-2


R.T.: 7.650 min
Delta R.T.: -0.004 min
Response: 947018
Conc: 301.98 ng/ml



#43 AR-1268-3


R.T.: 9.157 min
Delta R.T.: -0.002 min
Response: 47069
Conc: 8.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: -0.002 min
Response: 24594
Conc: 9.29 ng/ml




#44 AR-1268-4

R.T.: 9.598 min
Delta R.T.: -0.001 min
Response: 592046
Conc: 250.71 ng/ml



#44 AR-1268-4


R.T.: 8.146 min
Delta R.T.: -0.002 min
Response: 282777
Conc: 247.55 ng/ml



#45 AR-1268-5

R.T.: 10.032 min
Delta R.T.: -0.005 min
Response: 195631
Conc: 10.87 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.440 min
Delta R.T.: -0.002 min
Response: 92056
Conc: 10.54 ng/ml