

Data Path : Z:\PESTPCBSRV\HPCHEM1\ECD 0\Data\P0060318\
 Data File : P0045358.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jun 2018 16:25
 Operator : SM/UA
 Sample : PB109811BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB109811BS

Manual Integrations
 APPROVED

UGO
 6/6/2018 1:49:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 04:16:17 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0053018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 31 04:53:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.998	3.273	4170164	5608081	19.965	20.225
2)	SA Decachlor...	9.279	7.995	4371917	5438018	19.036	19.295
Target Compounds							
3)	L1 AR-1016-1	4.855	4.079	291536	418050	55.204m	60.492
4)	L1 AR-1016-2	5.193	4.508	436590	245303	65.427	41.601 #
5)	L1 AR-1016-3	5.287	4.545	339403	248450	60.657m	35.464 #
6)	L1 AR-1016-4	5.703	4.709	237447	946937	40.893m	127.195 #
7)	L1 AR-1016-5	5.935	5.063	4898052	2688635	974.746	358.592m#
31)	L7 AR-1260-1	6.655	5.687	694185	1043424	59.931	63.748
32)	L7 AR-1260-2	6.905	5.876	998567	1907162	65.623	90.269 #
33)	L7 AR-1260-3	7.255	6.218	676090	1015675	53.254	53.662
34)	L7 AR-1260-4	7.777	6.714	1478012	2250872	52.027	57.171
35)	L7 AR-1260-5	8.061	7.045	841828	1544961	50.320	56.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\PESTPCBSRV\HPCHEM1\ECD 0\Data\PO060318\
Data File : PO045358.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jun 2018 16:25
Operator : SM/UA
Sample : PB109811BS
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 04:16:17 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO053018.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 31 04:53:22 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.998 min
Delta R.T.: -0.002 min
Response: 4170164
Conc: 19.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#1 Tetrachloro-m-xylene

R.T.: 3.273 min
Delta R.T.: -0.002 min
Response: 5608081
Conc: 20.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.279 min
Delta R.T.: -0.004 min
Response: 4371917
Conc: 19.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.995 min
Delta R.T.: -0.003 min
Response: 5438018
Conc: 19.29 ng/ml




#3 AR-1016-1

R.T.: 4.855 min
Delta R.T.: -0.003 min
Response: 291536
Conc: 55.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#3 AR-1016-1

R.T.: 4.079 min
Delta R.T.: 0.000 min
Response: 418050
Conc: 60.49 ng/ml




#4 AR-1016-2

R.T.: 5.193 min
Delta R.T.: -0.001 min
Response: 436590
Conc: 65.43 ng/ml



#4 AR-1016-2

R.T.: 4.508 min
Delta R.T.: -0.002 min
Response: 245303
Conc: 41.60 ng/ml




#5 AR-1016-3

R.T.: 5.287 min
Delta R.T.: -0.002 min
Response: 339403
Conc: 60.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#5 AR-1016-3

R.T.: 4.545 min
Delta R.T.: -0.002 min
Response: 248450
Conc: 35.46 ng/ml




#6 AR-1016-4

R.T.: 5.703 min
Delta R.T.: -0.003 min
Response: 237447
Conc: 40.89 ng/ml m



#6 AR-1016-4

R.T.: 4.709 min
Delta R.T.: -0.001 min
Response: 946937
Conc: 127.20 ng/ml




#7 AR-1016-5

R.T.: 5.935 min
Delta R.T.: 0.002 min
Response: 4898052
Conc: 974.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#7 AR-1016-5

R.T.: 5.063 min
Delta R.T.: 0.022 min
Response: 2688635
Conc: 358.59 ng/ml m




#31 AR-1260-1

R.T.: 6.655 min
Delta R.T.: -0.004 min
Response: 694185
Conc: 59.93 ng/ml



#31 AR-1260-1

R.T.: 5.687 min
Delta R.T.: -0.003 min
Response: 1043424
Conc: 63.75 ng/ml




#32 AR-1260-2

R.T.: 6.905 min
Delta R.T.: -0.004 min
Response: 998567
Conc: 65.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#32 AR-1260-2

R.T.: 5.876 min
Delta R.T.: 0.000 min
Response: 1907162
Conc: 90.27 ng/ml




#33 AR-1260-3

R.T.: 7.255 min
Delta R.T.: -0.004 min
Response: 676090
Conc: 53.25 ng/ml



#33 AR-1260-3

R.T.: 6.218 min
Delta R.T.: -0.004 min
Response: 1015675
Conc: 53.66 ng/ml




#34 AR-1260-4

R.T.: 7.777 min
Delta R.T.: -0.003 min
Response: 1478012
Conc: 52.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB109811BS


Manual Integrations
APPROVED

UGO
6/6/2018 1:49:22 PM




#34 AR-1260-4

R.T.: 6.714 min
Delta R.T.: -0.002 min
Response: 2250872
Conc: 57.17 ng/ml



#35 AR-1260-5

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 841828
Conc: 50.32 ng/ml



#35 AR-1260-5

R.T.: 7.045 min
Delta R.T.: -0.002 min
Response: 1544961
Conc: 56.13 ng/ml