

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060319\
 Data File : P0056766.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Jun 2019 12:37
 Operator : SM/SJ
 Sample : K3085-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TANK-3

Manual Integrations
 APPROVED

Ankita
 6/4/2019 2:28:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 06:40:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.251	3.568	622804	682401	16.268m	20.692m#
2) SA Decachlor...	9.826	8.503	782944	448950	15.951m	12.591

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060319\
Data File : PO056766.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Jun 2019 12:37
Operator : SM/SJ
Sample : K3085-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TANK-3


Manual Integrations
APPROVED

Ankita
6/4/2019 2:28:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 06:40:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.251 min
Delta R.T.: -0.005 min
Response: 622804
Conc: 16.27 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TANK-3


Manual Integrations
APPROVED

Ankita
6/4/2019 2:28:06 PM




#1 Tetrachloro-m-xylene

R.T.: 3.568 min
Delta R.T.: -0.005 min
Response: 682401
Conc: 20.69 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.826 min
Delta R.T.: -0.010 min
Response: 782944
Conc: 15.95 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.503 min
Delta R.T.: -0.010 min
Response: 448950
Conc: 12.59 ng/ml