

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0060419\
 Data File : P0056817.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jun 2019 9:44
 Operator : SM/SJ
 Sample : PB120220BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB120220BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 11:13:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.249	3.567	793842	668945	20.736	20.284
2)	SA Decachlor...	9.821	8.500	1163883	740567	23.712	20.769
Target Compounds							
3)	L1 AR-1016-1	5.410	4.635	365761	251717	203.357	200.015
4)	L1 AR-1016-2	5.431	4.653	535484	359487	213.249	199.535
5)	L1 AR-1016-3	5.492	4.828	337513	192426	209.988	195.169
6)	L1 AR-1016-4	5.591	4.868	273589	165283	208.560	199.655
7)	L1 AR-1016-5	5.881	5.079	288736	167032	208.985	155.709 #
8)	L2 AR-1221-1	0.000	3.777	0	20538	N.D.	40.496 #
9)	L2 AR-1221-2	4.621	3.862	160777	32673	336.990	86.695 #
10)	L2 AR-1221-3	4.621	3.936	160777	123192	108.775	102.913
11)	L3 AR-1232-1	4.621	3.936	160777	123192	136.438	137.129
12)	L3 AR-1232-2	5.144	4.653	245732	359487	370.614	369.889
13)	L3 AR-1232-3	5.431	4.828	535484	192426	377.460	367.721
14)	L3 AR-1232-4	5.591	4.909	273589	197861	388.660	401.595
15)	L3 AR-1232-5	5.679	5.079	257540	167032	455.510	296.548 #
16)	L4 AR-1242-1	5.410	4.635	365761	251717	212.012	203.230
17)	L4 AR-1242-2	5.431	4.653	535484	359487	210.922	199.890
18)	L4 AR-1242-3	5.492	4.828	337513	192426	207.571	195.732
19)	L4 AR-1242-4	5.591	4.909	273589	197861	204.671	203.331
20)	L4 AR-1242-5	0.000	5.426	0	160298	N.D.	135.585 #
21)	L5 AR-1248-1	5.410	4.635	365761	251717	283.863	268.779
22)	L5 AR-1248-2	5.679	4.868	257540	165283	138.247	130.098
23)	L5 AR-1248-3	5.881	4.909	288736	197861	140.754	151.513
24)	L5 AR-1248-4	0.000	5.079	0	167032	N.D.	105.474 #
25)	L5 AR-1248-5	0.000	5.468	0	26434	N.D.	15.971 #
26)	L6 AR-1254-1	0.000	5.426	0	160298	N.D.	60.610 #
27)	L6 AR-1254-2	6.471	5.572	233357	144463	60.080	67.011
28)	L6 AR-1254-3	6.837	5.985	137521	300454	33.351	82.067 #
29)	L6 AR-1254-4	7.120	6.200	98883	38848	32.853	17.683 #
30)	L6 AR-1254-5	7.537	6.612	870352	515776	231.130	162.693 #
31)	L7 AR-1260-1	6.997	6.100	574713	415206	223.957	207.563
32)	L7 AR-1260-2	7.254	6.288	690209	511887	228.231	207.665
33)	L7 AR-1260-3	7.610	6.439	565660	467771	234.806	208.526
34)	L7 AR-1260-4	7.836	6.907	605904	407132	231.331	218.768
35)	L7 AR-1260-5	8.143	7.148	1116786	930832	233.535	215.854
36)	L8 AR-1262-1	7.610	6.649	565660	486962	120.947	116.664
37)	L8 AR-1262-2	8.143	7.148	1116786	930832	146.487	135.480
38)	L8 AR-1262-3	8.451	7.428	737536	235721	140.925	79.813 #
39)	L8 AR-1262-4	8.503	7.491	505956	658765	118.516	126.428
40)	L8 AR-1262-5	9.135	7.984	313551	232254	114.116	100.017
41)	L9 AR-1268-1	8.451	7.428	737536	235721	90.634	31.576 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060419\
 Data File : P0056817.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jun 2019 9:44
 Operator : SM/SJ
 Sample : PB120220BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB120220BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 11:13:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.503	7.491	505956	658765	62.542	94.337 #
44) L9	AR-1268-4	9.135	7.984	313551	232254	114.977	99.499
45) L9	AR-1268-5	9.514	8.260	92309	56001	4.404	3.236 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060419\
Data File : P0056817.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Jun 2019 9:44
Operator : SM/SJ
Sample : PB120220BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB120220BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 11:13:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.249 min
Delta R.T.: -0.007 min
Response: 793842
Conc: 20.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: -0.006 min
Response: 668945
Conc: 20.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.821 min
Delta R.T.: -0.015 min
Response: 1163883
Conc: 23.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.500 min
Delta R.T.: -0.013 min
Response: 740567
Conc: 20.77 ng/ml



#3 AR-1016-1


R.T.: 5.410 min
Delta R.T.: -0.008 min
Response: 365761
Conc: 203.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#3 AR-1016-1

R.T.: 4.635 min
Delta R.T.: -0.008 min
Response: 251717
Conc: 200.02 ng/ml




#4 AR-1016-2

R.T.: 5.431 min
Delta R.T.: -0.008 min
Response: 535484
Conc: 213.25 ng/ml



#4 AR-1016-2


R.T.: 4.653 min
Delta R.T.: -0.009 min
Response: 359487
Conc: 199.54 ng/ml



#5 AR-1016-3


R.T.: 5.492 min
Delta R.T.: -0.008 min
Response: 337513
Conc: 209.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#5 AR-1016-3

R.T.: 4.828 min
Delta R.T.: -0.008 min
Response: 192426
Conc: 195.17 ng/ml




#6 AR-1016-4

R.T.: 5.591 min
Delta R.T.: -0.008 min
Response: 273589
Conc: 208.56 ng/ml



#6 AR-1016-4


R.T.: 4.868 min
Delta R.T.: -0.008 min
Response: 165283
Conc: 199.65 ng/ml



#7 AR-1016-5


R.T.: 5.881 min
Delta R.T.: -0.009 min
Response: 288736
Conc: 208.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#7 AR-1016-5

R.T.: 5.079 min
Delta R.T.: -0.009 min
Response: 167032
Conc: 155.71 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.466 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.777 min
Delta R.T.: -0.007 min
Response: 20538
Conc: 40.50 ng/ml



#9 AR-1221-2


R.T.: 4.621 min
Delta R.T.: 0.069 min
Response: 160777
Conc: 336.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#9 AR-1221-2

R.T.: 3.862 min
Delta R.T.: -0.006 min
Response: 32673
Conc: 86.69 ng/ml




#10 AR-1221-3

R.T.: 4.621 min
Delta R.T.: -0.007 min
Response: 160777
Conc: 108.77 ng/ml



#10 AR-1221-3


R.T.: 3.936 min
Delta R.T.: -0.007 min
Response: 123192
Conc: 102.91 ng/ml



#11 AR-1232-1


R.T.: 4.621 min
Delta R.T.: -0.007 min
Response: 160777
Conc: 136.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#11 AR-1232-1

R.T.: 3.936 min
Delta R.T.: -0.007 min
Response: 123192
Conc: 137.13 ng/ml




#12 AR-1232-2

R.T.: 5.144 min
Delta R.T.: -0.008 min
Response: 245732
Conc: 370.61 ng/ml



#12 AR-1232-2


R.T.: 4.653 min
Delta R.T.: -0.008 min
Response: 359487
Conc: 369.89 ng/ml



#13 AR-1232-3


R.T.: 5.431 min
Delta R.T.: -0.008 min
Response: 535484
Conc: 377.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#13 AR-1232-3

R.T.: 4.828 min
Delta R.T.: -0.008 min
Response: 192426
Conc: 367.72 ng/ml




#14 AR-1232-4

R.T.: 5.591 min
Delta R.T.: -0.008 min
Response: 273589
Conc: 388.66 ng/ml



#14 AR-1232-4


R.T.: 4.909 min
Delta R.T.: -0.009 min
Response: 197861
Conc: 401.60 ng/ml



#15 AR-1232-5


R.T.: 5.679 min
Delta R.T.: -0.009 min
Response: 257540
Conc: 455.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#15 AR-1232-5

R.T.: 5.079 min
Delta R.T.: -0.009 min
Response: 167032
Conc: 296.55 ng/ml




#16 AR-1242-1

R.T.: 5.410 min
Delta R.T.: -0.008 min
Response: 365761
Conc: 212.01 ng/ml



#16 AR-1242-1


R.T.: 4.635 min
Delta R.T.: -0.007 min
Response: 251717
Conc: 203.23 ng/ml



#17 AR-1242-2


R.T.: 5.431 min
Delta R.T.: -0.008 min
Response: 535484
Conc: 210.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#17 AR-1242-2

R.T.: 4.653 min
Delta R.T.: -0.008 min
Response: 359487
Conc: 199.89 ng/ml




#18 AR-1242-3

R.T.: 5.492 min
Delta R.T.: -0.009 min
Response: 337513
Conc: 207.57 ng/ml



#18 AR-1242-3


R.T.: 4.828 min
Delta R.T.: -0.008 min
Response: 192426
Conc: 195.73 ng/ml



#19 AR-1242-4


R.T.: 5.591 min
Delta R.T.: -0.008 min
Response: 273589
Conc: 204.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#19 AR-1242-4

R.T.: 4.909 min
Delta R.T.: -0.008 min
Response: 197861
Conc: 203.33 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.329 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.426 min
Delta R.T.: -0.009 min
Response: 160298
Conc: 135.58 ng/ml



#21 AR-1248-1


R.T.: 5.410 min
Delta R.T.: -0.008 min
Response: 365761
Conc: 283.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#21 AR-1248-1

R.T.: 4.635 min
Delta R.T.: -0.007 min
Response: 251717
Conc: 268.78 ng/ml




#22 AR-1248-2

R.T.: 5.679 min
Delta R.T.: -0.009 min
Response: 257540
Conc: 138.25 ng/ml



#22 AR-1248-2


R.T.: 4.868 min
Delta R.T.: -0.007 min
Response: 165283
Conc: 130.10 ng/ml



#23 AR-1248-3


R.T.: 5.881 min
Delta R.T.: -0.008 min
Response: 288736
Conc: 140.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#23 AR-1248-3

R.T.: 4.909 min
Delta R.T.: -0.008 min
Response: 197861
Conc: 151.51 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.291 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.079 min
Delta R.T.: -0.008 min
Response: 167032
Conc: 105.47 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.329 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#25 AR-1248-5

R.T.: 5.468 min
Delta R.T.: -0.008 min
Response: 26434
Conc: 15.97 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.264 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.426 min
Delta R.T.: -0.009 min
Response: 160298
Conc: 60.61 ng/ml



#27 AR-1254-2


R.T.: 6.471 min
Delta R.T.: -0.010 min
Response: 233357
Conc: 60.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#27 AR-1254-2

R.T.: 5.572 min
Delta R.T.: -0.008 min
Response: 144463
Conc: 67.01 ng/ml




#28 AR-1254-3

R.T.: 6.837 min
Delta R.T.: -0.009 min
Response: 137521
Conc: 33.35 ng/ml



#28 AR-1254-3


R.T.: 5.985 min
Delta R.T.: 0.004 min
Response: 300454
Conc: 82.07 ng/ml



#29 AR-1254-4


R.T.: 7.120 min
Delta R.T.: -0.011 min
Response: 98883
Conc: 32.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#29 AR-1254-4

R.T.: 6.200 min
Delta R.T.: -0.009 min
Response: 38848
Conc: 17.68 ng/ml




#30 AR-1254-5

R.T.: 7.537 min
Delta R.T.: -0.010 min
Response: 870352
Conc: 231.13 ng/ml



#30 AR-1254-5


R.T.: 6.612 min
Delta R.T.: -0.010 min
Response: 515776
Conc: 162.69 ng/ml



#31 AR-1260-1


R.T.: 6.997 min
Delta R.T.: -0.010 min
Response: 574713
Conc: 223.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#31 AR-1260-1

R.T.: 6.100 min
Delta R.T.: -0.010 min
Response: 415206
Conc: 207.56 ng/ml




#32 AR-1260-2

R.T.: 7.254 min
Delta R.T.: -0.010 min
Response: 690209
Conc: 228.23 ng/ml



#32 AR-1260-2


R.T.: 6.288 min
Delta R.T.: -0.010 min
Response: 511887
Conc: 207.67 ng/ml



#33 AR-1260-3


R.T.: 7.610 min
Delta R.T.: -0.010 min
Response: 565660
Conc: 234.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#33 AR-1260-3

R.T.: 6.439 min
Delta R.T.: -0.011 min
Response: 467771
Conc: 208.53 ng/ml




#34 AR-1260-4

R.T.: 7.836 min
Delta R.T.: -0.009 min
Response: 605904
Conc: 231.33 ng/ml



#34 AR-1260-4


R.T.: 6.907 min
Delta R.T.: -0.010 min
Response: 407132
Conc: 218.77 ng/ml



#35 AR-1260-5


R.T.: 8.143 min
Delta R.T.: -0.011 min
Response: 1116786
Conc: 233.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#35 AR-1260-5

R.T.: 7.148 min
Delta R.T.: -0.010 min
Response: 930832
Conc: 215.85 ng/ml




#36 AR-1262-1

R.T.: 7.610 min
Delta R.T.: -0.010 min
Response: 565660
Conc: 120.95 ng/ml



#36 AR-1262-1


R.T.: 6.649 min
Delta R.T.: -0.011 min
Response: 486962
Conc: 116.66 ng/ml



#37 AR-1262-2


R.T.: 8.143 min
Delta R.T.: -0.011 min
Response: 1116786
Conc: 146.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#37 AR-1262-2

R.T.: 7.148 min
Delta R.T.: -0.011 min
Response: 930832
Conc: 135.48 ng/ml




#38 AR-1262-3

R.T.: 8.451 min
Delta R.T.: 0.006 min
Response: 737536
Conc: 140.92 ng/ml



#38 AR-1262-3


R.T.: 7.428 min
Delta R.T.: -0.011 min
Response: 235721
Conc: 79.81 ng/ml



#39 AR-1262-4


R.T.: 8.503 min
Delta R.T.: -0.028 min
Response: 505956
Conc: 118.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#39 AR-1262-4

R.T.: 7.491 min
Delta R.T.: -0.012 min
Response: 658765
Conc: 126.43 ng/ml




#40 AR-1262-5

R.T.: 9.135 min
Delta R.T.: -0.013 min
Response: 313551
Conc: 114.12 ng/ml



#40 AR-1262-5


R.T.: 7.984 min
Delta R.T.: -0.011 min
Response: 232254
Conc: 100.02 ng/ml



#41 AR-1268-1


R.T.: 8.451 min
Delta R.T.: 0.007 min
Response: 737536
Conc: 90.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#41 AR-1268-1

R.T.: 7.428 min
Delta R.T.: -0.010 min
Response: 235721
Conc: 31.58 ng/ml




#42 AR-1268-2

R.T.: 8.503 min
Delta R.T.: -0.031 min
Response: 505956
Conc: 62.54 ng/ml



#42 AR-1268-2


R.T.: 7.491 min
Delta R.T.: -0.012 min
Response: 658765
Conc: 94.34 ng/ml



#44 AR-1268-4


R.T.: 9.135 min
Delta R.T.: -0.013 min
Response: 313551
Conc: 114.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB120220BS




#44 AR-1268-4

R.T.: 7.984 min
Delta R.T.: -0.010 min
Response: 232254
Conc: 99.50 ng/ml



#45 AR-1268-5

R.T.: 9.514 min
Delta R.T.: -0.015 min
Response: 92309
Conc: 4.40 ng/ml



#45 AR-1268-5

R.T.: 8.260 min
Delta R.T.: -0.012 min
Response: 56001
Conc: 3.24 ng/ml