

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060419\
 Data File : P0056825.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jun 2019 13:49
 Operator : SM/SJ
 Sample : K3165-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 13911

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 14:06:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.257	3.569	579297	530228	15.132	16.078
2)	SA Decachlor...	9.832	8.503	448676	235606	9.141	6.608 #
Target Compounds							
3)	L1 AR-1016-1	5.437	4.637	125777	38960	69.930	30.958 #
4)	L1 AR-1016-2	5.437	4.655	125777	45548	50.089	25.282 #
5)	L1 AR-1016-3	5.437	4.831	125777	19335	78.254	19.611 #
6)	L1 AR-1016-4	0.000	4.831	0	19335	N.D.	23.356 #
7)	L1 AR-1016-5	0.000	5.144	0	38630	N.D.	36.011 #
8)	L2 AR-1221-1	4.434	0.000	662826	0	1049.225	N.D. #
10)	L2 AR-1221-3	0.000	3.963	0	101467	N.D.	84.764 #
11)	L3 AR-1232-1	0.000	3.963	0	101467	N.D.	112.946 #
12)	L3 AR-1232-2	5.152	4.655	44679	45548	67.384	46.866 #
13)	L3 AR-1232-3	5.437	4.831	125777	19335	88.660	36.949 #
14)	L3 AR-1232-4	0.000	4.938	0	43059	N.D.	87.397 #
15)	L3 AR-1232-5	5.693	5.144	71146	38630	125.836	68.583 #
16)	L4 AR-1242-1	5.437	4.637	125777	38960	72.906	31.456 #
17)	L4 AR-1242-2	5.437	4.655	125777	45548	49.542	25.327 #
18)	L4 AR-1242-3	5.437	4.831	125777	19335	77.353	19.667 #
19)	L4 AR-1242-4	0.000	4.938	0	43059	N.D.	44.250 #
20)	L4 AR-1242-5	0.000	5.434	0	31427	N.D.	26.582 #
21)	L5 AR-1248-1	5.437	4.637	125777	38960	97.614	41.601 #
22)	L5 AR-1248-2	5.693	4.831	71146	19335	38.191	15.219 #
23)	L5 AR-1248-3	0.000	4.938	0	43059	N.D.	32.973 #
24)	L5 AR-1248-4	0.000	5.144	0	38630	N.D.	24.393 #
25)	L5 AR-1248-5	0.000	5.434	0	31427	N.D.	18.988 #
26)	L6 AR-1254-1	0.000	5.434	0	31427	N.D.	11.883 #
27)	L6 AR-1254-2	6.483	5.578	48426	10801	12.468	5.010 #
28)	L6 AR-1254-3	6.808	5.985	190659	41214	46.238	11.257 #
29)	L6 AR-1254-4	7.130	6.165	32730	15415	10.874	7.017 #
30)	L6 AR-1254-5	7.545	6.614	162123	86838	43.053	27.392 #
31)	L7 AR-1260-1	7.028	6.103	1296119	64343	505.078	32.165 #
32)	L7 AR-1260-2	7.254	6.297	182426	107926	60.323	43.784 #
33)	L7 AR-1260-3	7.618	6.443	117554	54969	48.797	24.504 #
34)	L7 AR-1260-4	7.844	6.909	150539	34083	57.475	18.314 #
35)	L7 AR-1260-5	8.149	7.151	337319	135947	70.538	31.525 #
36)	L8 AR-1262-1	7.618	6.653	117554	86488	25.135	20.720
37)	L8 AR-1262-2	8.149	7.151	337319	135947	44.246	19.787 #
38)	L8 AR-1262-3	8.454	7.430	308441	98753	58.935	33.437 #
39)	L8 AR-1262-4	8.511	7.483	303282	933481	71.041	179.150 #
40)	L8 AR-1262-5	9.146	7.993	151485	126251	55.133	54.368
41)	L9 AR-1268-1	8.454	7.430	308441	98753	37.904	13.229 #
42)	L9 AR-1268-2	8.511	7.483	303282	933481	37.489	133.678 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060419\
 Data File : P0056825.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jun 2019 13:49
 Operator : SM/SJ
 Sample : K3165-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 13911

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 14:06:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.760	7.722	2053839	25920	311.706	4.367 #
44) L9	AR-1268-4	9.146	7.993	151485	126251	55.549	54.086
45) L9	AR-1268-5	0.000	8.262	0	24582	N.D.	1.420 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060419\
Data File : P0056825.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Jun 2019 13:49
Operator : SM/SJ
Sample : K3165-06
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
13911

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 14:06:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.257 min
Delta R.T.: 0.000 min
Response: 579297
Conc: 15.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#1 Tetrachloro-m-xylene

R.T.: 3.569 min
Delta R.T.: -0.004 min
Response: 530228
Conc: 16.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.832 min
Delta R.T.: -0.005 min
Response: 448676
Conc: 9.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.503 min
Delta R.T.: -0.010 min
Response: 235606
Conc: 6.61 ng/ml



#3 AR-1016-1


R.T.: 5.437 min
Delta R.T.: 0.019 min
Response: 125777
Conc: 69.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#3 AR-1016-1

R.T.: 4.637 min
Delta R.T.: -0.006 min
Response: 38960
Conc: 30.96 ng/ml




#4 AR-1016-2

R.T.: 5.437 min
Delta R.T.: -0.002 min
Response: 125777
Conc: 50.09 ng/ml



#4 AR-1016-2


R.T.: 4.655 min
Delta R.T.: -0.006 min
Response: 45548
Conc: 25.28 ng/ml



#5 AR-1016-3


R.T.: 5.437 min
Delta R.T.: -0.063 min
Response: 125777
Conc: 78.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#5 AR-1016-3

R.T.: 4.831 min
Delta R.T.: -0.005 min
Response: 19335
Conc: 19.61 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.599 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 4.831 min
Delta R.T.: -0.046 min
Response: 19335
Conc: 23.36 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 5.890 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#7 AR-1016-5

R.T.: 5.144 min
Delta R.T.: 0.057 min
Response: 38630
Conc: 36.01 ng/ml




#8 AR-1221-1

R.T.: 4.434 min
Delta R.T.: -0.032 min
Response: 662826
Conc: 1049.22 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.783 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.628 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#10 AR-1221-3

R.T.: 3.963 min
Delta R.T.: 0.020 min
Response: 101467
Conc: 84.76 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.629 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 3.963 min
Delta R.T.: 0.020 min
Response: 101467
Conc: 112.95 ng/ml



#12 AR-1232-2


R.T.: 5.152 min
Delta R.T.: 0.000 min
Response: 44679
Conc: 67.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#12 AR-1232-2

R.T.: 4.655 min
Delta R.T.: -0.006 min
Response: 45548
Conc: 46.87 ng/ml




#13 AR-1232-3

R.T.: 5.437 min
Delta R.T.: -0.002 min
Response: 125777
Conc: 88.66 ng/ml



#13 AR-1232-3


R.T.: 4.831 min
Delta R.T.: -0.005 min
Response: 19335
Conc: 36.95 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.599 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#14 AR-1232-4

R.T.: 4.938 min
Delta R.T.: 0.020 min
Response: 43059
Conc: 87.40 ng/ml




#15 AR-1232-5

R.T.: 5.693 min
Delta R.T.: 0.004 min
Response: 71146
Conc: 125.84 ng/ml



#15 AR-1232-5


R.T.: 5.144 min
Delta R.T.: 0.056 min
Response: 38630
Conc: 68.58 ng/ml



#16 AR-1242-1


R.T.: 5.437 min
Delta R.T.: 0.019 min
Response: 125777
Conc: 72.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#16 AR-1242-1

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 38960
Conc: 31.46 ng/ml




#17 AR-1242-2

R.T.: 5.437 min
Delta R.T.: -0.002 min
Response: 125777
Conc: 49.54 ng/ml



#17 AR-1242-2


R.T.: 4.655 min
Delta R.T.: -0.006 min
Response: 45548
Conc: 25.33 ng/ml



#18 AR-1242-3


R.T.: 5.437 min
Delta R.T.: -0.063 min
Response: 125777
Conc: 77.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#18 AR-1242-3

R.T.: 4.831 min
Delta R.T.: -0.004 min
Response: 19335
Conc: 19.67 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.599 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 4.938 min
Delta R.T.: 0.021 min
Response: 43059
Conc: 44.25 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.329 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#20 AR-1242-5

R.T.: 5.434 min
Delta R.T.: -0.001 min
Response: 31427
Conc: 26.58 ng/ml




#21 AR-1248-1

R.T.: 5.437 min
Delta R.T.: 0.020 min
Response: 125777
Conc: 97.61 ng/ml



#21 AR-1248-1


R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 38960
Conc: 41.60 ng/ml



#22 AR-1248-2


R.T.: 5.693 min
Delta R.T.: 0.005 min
Response: 71146
Conc: 38.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#22 AR-1248-2

R.T.: 4.831 min
Delta R.T.: -0.045 min
Response: 19335
Conc: 15.22 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.890 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 4.938 min
Delta R.T.: 0.021 min
Response: 43059
Conc: 32.97 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 6.291 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#24 AR-1248-4

R.T.: 5.144 min
Delta R.T.: 0.057 min
Response: 38630
Conc: 24.39 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.329 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.434 min
Delta R.T.: -0.042 min
Response: 31427
Conc: 18.99 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.264 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
13911




#26 AR-1254-1

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 31427
Conc: 11.88 ng/ml




#27 AR-1254-2

R.T.: 6.483 min
Delta R.T.: 0.002 min
Response: 48426
Conc: 12.47 ng/ml



#27 AR-1254-2


R.T.: 5.578 min
Delta R.T.: -0.002 min
Response: 10801
Conc: 5.01 ng/ml



#28 AR-1254-3


R.T.: 6.808 min
Delta R.T.: -0.038 min
Response: 190659
Conc: 46.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#28 AR-1254-3

R.T.: 5.985 min
Delta R.T.: 0.003 min
Response: 41214
Conc: 11.26 ng/ml




#29 AR-1254-4

R.T.: 7.130 min
Delta R.T.: -0.002 min
Response: 32730
Conc: 10.87 ng/ml



#29 AR-1254-4


R.T.: 6.165 min
Delta R.T.: -0.043 min
Response: 15415
Conc: 7.02 ng/ml



#30 AR-1254-5


R.T.: 7.545 min
Delta R.T.: -0.002 min
Response: 162123
Conc: 43.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#30 AR-1254-5

R.T.: 6.614 min
Delta R.T.: -0.008 min
Response: 86838
Conc: 27.39 ng/ml




#31 AR-1260-1

R.T.: 7.028 min
Delta R.T.: 0.021 min
Response: 1296119
Conc: 505.08 ng/ml



#31 AR-1260-1


R.T.: 6.103 min
Delta R.T.: -0.007 min
Response: 64343
Conc: 32.17 ng/ml



#32 AR-1260-2


R.T.: 7.254 min
Delta R.T.: -0.010 min
Response: 182426
Conc: 60.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#32 AR-1260-2

R.T.: 6.297 min
Delta R.T.: 0.000 min
Response: 107926
Conc: 43.78 ng/ml




#33 AR-1260-3

R.T.: 7.618 min
Delta R.T.: -0.002 min
Response: 117554
Conc: 48.80 ng/ml



#33 AR-1260-3


R.T.: 6.443 min
Delta R.T.: -0.007 min
Response: 54969
Conc: 24.50 ng/ml



#34 AR-1260-4


R.T.: 7.844 min
Delta R.T.: -0.001 min
Response: 150539
Conc: 57.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#34 AR-1260-4

R.T.: 6.909 min
Delta R.T.: -0.008 min
Response: 34083
Conc: 18.31 ng/ml




#35 AR-1260-5

R.T.: 8.149 min
Delta R.T.: -0.005 min
Response: 337319
Conc: 70.54 ng/ml



#35 AR-1260-5


R.T.: 7.151 min
Delta R.T.: -0.007 min
Response: 135947
Conc: 31.53 ng/ml



#36 AR-1262-1


R.T.: 7.618 min
Delta R.T.: -0.003 min
Response: 117554
Conc: 25.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#36 AR-1262-1

R.T.: 6.653 min
Delta R.T.: -0.006 min
Response: 86488
Conc: 20.72 ng/ml




#37 AR-1262-2

R.T.: 8.149 min
Delta R.T.: -0.005 min
Response: 337319
Conc: 44.25 ng/ml



#37 AR-1262-2


R.T.: 7.151 min
Delta R.T.: -0.007 min
Response: 135947
Conc: 19.79 ng/ml



#38 AR-1262-3


R.T.: 8.454 min
Delta R.T.: 0.009 min
Response: 308441
Conc: 58.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#38 AR-1262-3

R.T.: 7.430 min
Delta R.T.: -0.008 min
Response: 98753
Conc: 33.44 ng/ml




#39 AR-1262-4

R.T.: 8.511 min
Delta R.T.: -0.019 min
Response: 303282
Conc: 71.04 ng/ml



#39 AR-1262-4


R.T.: 7.483 min
Delta R.T.: -0.020 min
Response: 933481
Conc: 179.15 ng/ml



#40 AR-1262-5


R.T.: 9.146 min
Delta R.T.: -0.002 min
Response: 151485
Conc: 55.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#40 AR-1262-5

R.T.: 7.993 min
Delta R.T.: -0.001 min
Response: 126251
Conc: 54.37 ng/ml




#41 AR-1268-1

R.T.: 8.454 min
Delta R.T.: 0.009 min
Response: 308441
Conc: 37.90 ng/ml



#41 AR-1268-1


R.T.: 7.430 min
Delta R.T.: -0.007 min
Response: 98753
Conc: 13.23 ng/ml



#42 AR-1268-2


R.T.: 8.511 min
Delta R.T.: -0.022 min
Response: 303282
Conc: 37.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#42 AR-1268-2

R.T.: 7.483 min
Delta R.T.: -0.020 min
Response: 933481
Conc: 133.68 ng/ml




#43 AR-1268-3

R.T.: 8.760 min
Delta R.T.: 0.017 min
Response: 2053839
Conc: 311.71 ng/ml



#43 AR-1268-3


R.T.: 7.722 min
Delta R.T.: 0.017 min
Response: 25920
Conc: 4.37 ng/ml



#44 AR-1268-4


R.T.: 9.146 min
Delta R.T.: -0.002 min
Response: 151485
Conc: 55.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
13911




#44 AR-1268-4

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 126251
Conc: 54.09 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 9.529 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.262 min
Delta R.T.: -0.010 min
Response: 24582
Conc: 1.42 ng/ml