

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060520\
 Data File : P0068786.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jun 2020 3:34
 Operator : DD\AJ
 Sample : L2890-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CHRT-26160

Manual Integrations
 APPROVED

Sohil
 6/8/2020 10:44:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 05:20:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060520.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 06 05:06:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.782	3.852	506411	672048	16.367m	17.752m
2) SA Decachlor...	10.662	9.101	1048717	1170160	24.497	24.027m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060520\
Data File : P0068786.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jun 2020 3:34
Operator : DD\AJ
Sample : L2890-01
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CHRT-26160


Manual Integrations
APPROVED

Sohil
6/8/2020 10:44:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 05:20:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060520.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 06 05:06:44 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.782 min
Delta R.T.: 0.000 min
Response: 506411
Conc: 16.37 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CHRT-26160


Manual Integrations
APPROVED

Sohil
6/8/2020 10:44:54 AM




#1 Tetrachloro-m-xylene

R.T.: 3.852 min
Delta R.T.: 0.001 min
Response: 672048
Conc: 17.75 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.662 min
Delta R.T.: 0.000 min
Response: 1048717
Conc: 24.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: 0.000 min
Response: 1170160
Conc: 24.03 ng/ml m