

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060520\
 Data File : P0068821.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jun 2020 17:02
 Operator : DD\AJ
 Sample : L2918-06DL 2X
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 6-19029-TC-1DL

Manual Integrations
 APPROVED

Sohil
 6/8/2020 10:45:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:22:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060520.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 06 05:06:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.779	3.849	373686	492184	12.078m	13.001
2)	SA Decachlor...	10.659	9.101	504296	574518	11.780	11.797
Target Compounds							
26)	L6 AR-1254-1	7.002	5.964	787994	1579376	620.697	639.123
27)	L6 AR-1254-2	7.230	6.124	1287213	1260520	632.389	564.586
28)	L6 AR-1254-3	7.610	6.540	1594624	2360013	750.109	692.379
29)	L6 AR-1254-4	7.906	6.779	1607891	1954053	830.516	816.580
30)	L6 AR-1254-5	8.332	7.205	1120977	1792260	599.319	570.359

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060520\
Data File : PO068821.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jun 2020 17:02
Operator : DD\AJ
Sample : L2918-06DL 2X
Misc :
ALS Vial : 63 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
6-19029-TC-1DL


Manual Integrations
APPROVED

Sohil
6/8/2020 10:45:07 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:22:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060520.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 06 05:06:44 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.779 min
Delta R.T.: -0.003 min
Response: 373686
Conc: 12.08 ng/ml m

Instrument :
ECD_O
ClientSampleId :
6-19029-TC-1DL


Manual Integrations
APPROVED

Sohil
6/8/2020 10:45:07 AM




#1 Tetrachloro-m-xylene

R.T.: 3.849 min
Delta R.T.: -0.002 min
Response: 492184
Conc: 13.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.659 min
Delta R.T.: -0.004 min
Response: 504296
Conc: 11.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: 0.000 min
Response: 574518
Conc: 11.80 ng/ml




#26 AR-1254-1

R.T.: 7.002 min
Delta R.T.: -0.002 min
Response: 787994
Conc: 620.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-19029-TC-1DL


Manual Integrations
APPROVED

Sohil
6/8/2020 10:45:07 AM




#26 AR-1254-1

R.T.: 5.964 min
Delta R.T.: 0.000 min
Response: 1579376
Conc: 639.12 ng/ml




#27 AR-1254-2

R.T.: 7.230 min
Delta R.T.: -0.002 min
Response: 1287213
Conc: 632.39 ng/ml



#27 AR-1254-2

R.T.: 6.124 min
Delta R.T.: -0.001 min
Response: 1260520
Conc: 564.59 ng/ml




#28 AR-1254-3

R.T.: 7.610 min
Delta R.T.: -0.002 min
Response: 1594624
Conc: 750.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-19029-TC-1DL


Manual Integrations
APPROVED

Sohil
6/8/2020 10:45:07 AM




#28 AR-1254-3

R.T.: 6.540 min
Delta R.T.: -0.001 min
Response: 2360013
Conc: 692.38 ng/ml




#29 AR-1254-4

R.T.: 7.906 min
Delta R.T.: -0.002 min
Response: 1607891
Conc: 830.52 ng/ml



#29 AR-1254-4

R.T.: 6.779 min
Delta R.T.: 0.000 min
Response: 1954053
Conc: 816.58 ng/ml




#30 AR-1254-5

R.T.: 8.332 min
Delta R.T.: 0.000 min
Response: 1120977
Conc: 599.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-19029-TC-1DL

Manual Integrations
APPROVED

Sohil
6/8/2020 10:45:07 AM



#30 AR-1254-5

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 1792260
Conc: 570.36 ng/ml