

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060719\
 Data File : P0057003.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2019 1:18
 Operator : SM/SJ
 Sample : K3204-02
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-2

Manual Integrations
 APPROVED

Ankita
 6/10/2019 12:30:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 04:07:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.243	3.564	240343	231028	6.278m	7.005
2) SA Decachlor...	9.812	8.494	163582	131194	3.333m	3.679

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060719\
Data File : PO057003.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jun 2019 1:18
Operator : SM/SJ
Sample : K3204-02
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
COMP-2


Manual Integrations
APPROVED

Ankita
6/10/2019 12:30:43 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 04:07:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.243 min
Delta R.T.: -0.013 min
Response: 240343
Conc: 6.28 ng/ml m

Instrument :
ECD_O
ClientSampleId :
COMP-2


Manual Integrations
APPROVED

Ankita
6/10/2019 12:30:43 PM




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: -0.009 min
Response: 231028
Conc: 7.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.812 min
Delta R.T.: -0.025 min
Response: 163582
Conc: 3.33 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.494 min
Delta R.T.: -0.019 min
Response: 131194
Conc: 3.68 ng/ml