

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061318\
Data File : P0045740.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jun 2018 18:41
Operator : SM/UA
Sample : J3251-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SU-03-061118

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:02:27 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061218.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 13 03:54:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.631	4499709	2423840	23.092m	21.324m
2) SA Decachlor...	10.060	8.615	2824868	1104181	20.177m	15.790

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061318\
Data File : P0045740.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jun 2018 18:41
Operator : SM/UA
Sample : J3251-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SU-03-061118

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:02:27 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061218.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 13 03:54:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.009 min
Response: 4499709
Conc: 23.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-03-061118




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.005 min
Response: 2423840
Conc: 21.32 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.060 min
Delta R.T.: 0.018 min
Response: 2824868
Conc: 20.18 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.615 min
Delta R.T.: 0.004 min
Response: 1104181
Conc: 15.79 ng/ml