

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061318\
 Data File : P0045741.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2018 18:57
 Operator : SM/UA
 Sample : J3459-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MH-RO-341

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:03:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061218.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 13 03:54:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.398	3.631	4103316	2291718	21.058m	20.162m
2) SA Decachlor...	10.061	8.615	3030434	1154918	21.645m	16.516m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061318\
Data File : P0045741.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jun 2018 18:57
Operator : SM/UA
Sample : J3459-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MH-RO-341

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:03:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061218.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 13 03:54:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.398 min
Delta R.T.: 0.010 min
Response: 4103316
Conc: 21.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-RO-341




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.005 min
Response: 2291718
Conc: 20.16 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.019 min
Response: 3030434
Conc: 21.65 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.615 min
Delta R.T.: 0.004 min
Response: 1154918
Conc: 16.52 ng/ml m