

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061920\
 Data File : P0069040.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jun 2020 9:33
 Operator : DD\AJ
 Sample : L3034-14
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-5

Manual Integrations
 APPROVED

Sohil
 6/22/2020 4:27:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 11:02:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 19 05:15:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.405	3.565	703200	949751	23.770	24.676
2)	SA Decachlor...	10.070	8.771	916131	998450	19.289	18.342
Target Compounds							
26)	L6 AR-1254-1	6.606	5.652	20705	35895	13.667m	15.267
27)	L6 AR-1254-2	6.840	5.814	50886	29395	21.216m	13.938m#
28)	L6 AR-1254-3	7.230	6.225	66882	39324	25.191m	11.404 #
29)	L6 AR-1254-4	7.508	6.464	54662	26450	25.633	11.343m#
30)	L6 AR-1254-5	7.931	6.888	93172	76296	41.567	23.565 #
31)	L7 AR-1260-1	7.377	6.361	61724	56739	32.343	24.248m#
32)	L7 AR-1260-2	7.646	6.561	137426	77174	58.716	27.111 #
33)	L7 AR-1260-3	8.004	6.709	45820	58760	25.792	21.847
34)	L7 AR-1260-4	8.230	7.190	45241	37306	21.694	15.571m#
35)	L7 AR-1260-5	8.545	7.440	96990	96411	22.066	16.642

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061920\
Data File : PO069040.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jun 2020 9:33
Operator : DD\AJ
Sample : L3034-14
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 11:02:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 19 05:15:02 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 703200
Conc: 23.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: 0.001 min
Response: 949751
Conc: 24.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.070 min
Delta R.T.: 0.000 min
Response: 916131
Conc: 19.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.771 min
Delta R.T.: 0.000 min
Response: 998450
Conc: 18.34 ng/ml




#26 AR-1254-1

R.T.: 6.606 min
Delta R.T.: -0.001 min
Response: 20705
Conc: 13.67 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#26 AR-1254-1

R.T.: 5.652 min
Delta R.T.: 0.000 min
Response: 35895
Conc: 15.27 ng/ml




#27 AR-1254-2

R.T.: 6.840 min
Delta R.T.: 0.004 min
Response: 50886
Conc: 21.22 ng/ml m



#27 AR-1254-2

R.T.: 5.814 min
Delta R.T.: 0.001 min
Response: 29395
Conc: 13.94 ng/ml m




#28 AR-1254-3

R.T.: 7.230 min
Delta R.T.: 0.017 min
Response: 66882
Conc: 25.19 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#28 AR-1254-3

R.T.: 6.225 min
Delta R.T.: 0.000 min
Response: 39324
Conc: 11.40 ng/ml




#29 AR-1254-4

R.T.: 7.508 min
Delta R.T.: 0.000 min
Response: 54662
Conc: 25.63 ng/ml



#29 AR-1254-4

R.T.: 6.464 min
Delta R.T.: 0.000 min
Response: 26450
Conc: 11.34 ng/ml m




#30 AR-1254-5

R.T.: 7.931 min
Delta R.T.: 0.000 min
Response: 93172
Conc: 41.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#30 AR-1254-5

R.T.: 6.888 min
Delta R.T.: 0.000 min
Response: 76296
Conc: 23.56 ng/ml




#31 AR-1260-1

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 61724
Conc: 32.34 ng/ml



#31 AR-1260-1

R.T.: 6.361 min
Delta R.T.: -0.001 min
Response: 56739
Conc: 24.25 ng/ml m




#32 AR-1260-2

R.T.: 7.646 min
Delta R.T.: 0.002 min
Response: 137426
Conc: 58.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#32 AR-1260-2

R.T.: 6.561 min
Delta R.T.: 0.000 min
Response: 77174
Conc: 27.11 ng/ml




#33 AR-1260-3

R.T.: 8.004 min
Delta R.T.: 0.000 min
Response: 45820
Conc: 25.79 ng/ml



#33 AR-1260-3

R.T.: 6.709 min
Delta R.T.: -0.002 min
Response: 58760
Conc: 21.85 ng/ml




#34 AR-1260-4

R.T.: 8.230 min
Delta R.T.: 0.000 min
Response: 45241
Conc: 21.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-5


Manual Integrations
APPROVED

Sohil
6/22/2020 4:27:27 PM




#34 AR-1260-4

R.T.: 7.190 min
Delta R.T.: 0.000 min
Response: 37306
Conc: 15.57 ng/ml m



#35 AR-1260-5

R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 96990
Conc: 22.07 ng/ml



#35 AR-1260-5

R.T.: 7.440 min
Delta R.T.: 0.000 min
Response: 96411
Conc: 16.64 ng/ml