

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0062018\
 Data File : P0045977.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jun 2018 14:57
 Operator : SM/UA
 Sample : J3551-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OILY-DEBRIS

Manual Integrations
 APPROVED

Sohil
 6/23/2018 1:36:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 15:26:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 20 03:39:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.400	3.635	6317103	2574685	13.210	10.242
2) SA Decachlor...	10.194f	8.736f	3966423	1229446	9.463m	7.850m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062018\
Data File : P0045977.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2018 14:57
Operator : SM/UA
Sample : J3551-03
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
OILY-DEBRIS


Manual Integrations
APPROVED

Sohil
6/23/2018 1:36:41 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 15:26:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061918.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 03:39:54 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.400 min
Delta R.T.: 0.016 min
Response: 6317103
Conc: 13.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
OILY-DEBRIS


Manual Integrations
APPROVED

Sohil
6/23/2018 1:36:41 AM




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: 0.023 min
Response: 2574685
Conc: 10.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.194 min
Delta R.T.: 0.140 min
Response: 3966423
Conc: 9.46 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.736 min
Delta R.T.: 0.143 min
Response: 1229446
Conc: 7.85 ng/ml m