

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062018\
 Data File : P0045989.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jun 2018 18:13
 Operator : SM/UA
 Sample : PB110231BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB110231BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:12:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 20 03:39:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.390	3.612	10789093	5529139	22.562	21.994
2) SA Decachlor...	10.067	8.593	6908606	2904714	16.482	18.547

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062018\
Data File : P0045989.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2018 18:13
Operator : SM/UA
Sample : PB110231BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB110231BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:12:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061918.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 03:39:54 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.390 min
Delta R.T.: 0.007 min
Response: 10789093
Conc: 22.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB110231BL




#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: 0.000 min
Response: 5529139
Conc: 21.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.067 min
Delta R.T.: 0.013 min
Response: 6908606
Conc: 16.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.593 min
Delta R.T.: 0.000 min
Response: 2904714
Conc: 18.55 ng/ml