

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062220\
 Data File : P0069115.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jun 2020 22:16
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 05:33:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 19 05:15:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.405	3.565	709445	863755	23.981	22.441
2) SA Decachlor...	10.071	8.773	991098	1115469	20.867	20.491
Target Compounds						
40) L8 AR-1262-5	9.468	0.000	24258	0	13.520	N.D. #
44) L9 AR-1268-4	9.468	0.000	24258	0	11.576	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062220\
Data File : P0069115.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jun 2020 22:16
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 05:33:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 19 05:15:02 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 709445
Conc: 23.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: 0.001 min
Response: 863755
Conc: 22.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.071 min
Delta R.T.: 0.001 min
Response: 991098
Conc: 20.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: 0.002 min
Response: 1115469
Conc: 20.49 ng/ml



#40 AR-1262-5


R.T.: 9.468 min
Delta R.T.: 0.008 min
Response: 24258
Conc: 13.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 8.283 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 9.468 min
Delta R.T.: 0.008 min
Response: 24258
Conc: 11.58 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.283 min
Response: 0
Conc: N.D.