

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0062419\
 Data File : P0057394.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jun 2019 12:30
 Operator : SM/SJ
 Sample : P0062419ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 ICVPO062419

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 14:04:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.230	3.555	1881965	1772367	47.910	50.404
2)	SA Decachlor...	9.783	8.476	2716503	1875311	48.308	50.133
Target Compounds							
3)	L1 AR-1016-1	5.390	4.621	830689	650638	469.990	486.647
4)	L1 AR-1016-2	5.412	4.639	1224287	955865	475.085	490.973
5)	L1 AR-1016-3	5.473	4.813	747902	522746	479.776	493.745
6)	L1 AR-1016-4	5.571	4.853	615588	434544	474.702	488.413
7)	L1 AR-1016-5	5.862	5.064	650864	561578	476.181	491.575
8)	L2 AR-1221-1	4.438	3.764	60136	57404	138.134	155.341
9)	L2 AR-1221-2	4.528	3.850	96571	87077	290.609	311.993
10)	L2 AR-1221-3	4.603	3.923	377440	322760	357.551	368.513
11)	L3 AR-1232-1	4.603	3.923	377440	322760	420.450	432.042
12)	L3 AR-1232-2	5.125	4.639	558193	955865	1151.894	1160.129
13)	L3 AR-1232-3	5.412	4.813	1224287	522746	1167.321	1210.240
14)	L3 AR-1232-4	5.571	4.894	615588	517871	1158.166	1362.091
15)	L3 AR-1232-5	5.660	5.064	563225	561578	1363.949	1319.067
16)	L4 AR-1242-1	5.390	4.621	830689	650638	586.129	614.796
17)	L4 AR-1242-2	5.412	4.639	1224287	955865	588.642	613.538
18)	L4 AR-1242-3	5.473	4.813	747902	522746	592.422	616.524
19)	L4 AR-1242-4	5.571	4.894	615588	517871	567.407	620.585
20)	L4 AR-1242-5	6.301	5.410	112253	429188	86.510	428.562 #
21)	L5 AR-1248-1	5.390	4.621	830689	650638	713.209	701.681
22)	L5 AR-1248-2	5.660	4.853	563225	434544	324.458	338.072
23)	L5 AR-1248-3	5.862	4.894	650864	517871	334.183	392.445
24)	L5 AR-1248-4	6.262	5.064	138729	561578	59.188	346.982 #
25)	L5 AR-1248-5	6.301	5.451	112253	72134	49.988	46.642
26)	L6 AR-1254-1	6.234	5.410	489217	429188	251.170	201.808
27)	L6 AR-1254-2	6.450	5.556	554351	384095	177.861	206.363
28)	L6 AR-1254-3	6.815	5.968	317633	789570	96.076	262.210 #
29)	L6 AR-1254-4	7.098	6.182	251310	104027	92.601	57.791 #
30)	L6 AR-1254-5	7.514	6.594	1926323	1345277	688.984	496.576 #
31)	L7 AR-1260-1	6.976	6.083	1327455	1050111	478.989	490.639
32)	L7 AR-1260-2	7.232	6.270	1722579	1329621	467.254	484.622
33)	L7 AR-1260-3	7.587	6.421	1463901	1202262	481.841	492.427
34)	L7 AR-1260-4	7.813	6.887	1446692	1031228	477.026	495.296
35)	L7 AR-1260-5	8.119	7.130	2986496	2510138	477.568	494.888
36)	L8 AR-1262-1	7.587	6.630	1463901	1272756	322.103	340.015
37)	L8 AR-1262-2	8.119	7.130	2986496	2510138	416.835	422.451
38)	L8 AR-1262-3	8.426	7.408	1867377	620249	380.313	242.491 #
39)	L8 AR-1262-4	8.477	7.472	1329106	1757204	326.697	398.699
40)	L8 AR-1262-5	9.104	7.965	868876	623171	344.562	341.229
41)	L9 AR-1268-1	8.426	7.408	1867377	620249	221.461	90.587 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062419\
 Data File : P0057394.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jun 2019 12:30
 Operator : SM/SJ
 Sample : P0062419ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1


Instrument :
 ECD_O
 ClientSampleId :
 ICVPO062419

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 14:04:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.477	7.472	1329106	1757204	166.462	276.852 #
43) L9	AR-1268-3	8.701	7.673	56242	40336	8.343	7.538
44) L9	AR-1268-4	9.104	7.965	868876	623171	337.219	311.059
45) L9	AR-1268-5	9.480	8.239	267816	170750	13.539	11.112


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 4.230 min
Delta R.T.: 0.000 min
Response: 1881965
Conc: 47.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#1 Tetrachloro-m-xylene

R.T.: 3.555 min
Delta R.T.: 0.000 min
Response: 1772367
Conc: 50.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.783 min
Delta R.T.: 0.001 min
Response: 2716503
Conc: 48.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: -0.001 min
Response: 1875311
Conc: 50.13 ng/ml



#3 AR-1016-1


R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 830689
Conc: 469.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#3 AR-1016-1

R.T.: 4.621 min
Delta R.T.: 0.000 min
Response: 650638
Conc: 486.65 ng/ml




#4 AR-1016-2

R.T.: 5.412 min
Delta R.T.: 0.000 min
Response: 1224287
Conc: 475.09 ng/ml



#4 AR-1016-2


R.T.: 4.639 min
Delta R.T.: 0.000 min
Response: 955865
Conc: 490.97 ng/ml



#5 AR-1016-3


R.T.: 5.473 min
Delta R.T.: 0.000 min
Response: 747902
Conc: 479.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#5 AR-1016-3

R.T.: 4.813 min
Delta R.T.: 0.000 min
Response: 522746
Conc: 493.74 ng/ml




#6 AR-1016-4

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 615588
Conc: 474.70 ng/ml



#6 AR-1016-4


R.T.: 4.853 min
Delta R.T.: -0.001 min
Response: 434544
Conc: 488.41 ng/ml



#7 AR-1016-5


R.T.: 5.862 min
Delta R.T.: 0.001 min
Response: 650864
Conc: 476.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#7 AR-1016-5

R.T.: 5.064 min
Delta R.T.: 0.000 min
Response: 561578
Conc: 491.58 ng/ml




#8 AR-1221-1

R.T.: 4.438 min
Delta R.T.: -0.011 min
Response: 60136
Conc: 138.13 ng/ml



#8 AR-1221-1


R.T.: 3.764 min
Delta R.T.: -0.003 min
Response: 57404
Conc: 155.34 ng/ml



#9 AR-1221-2


R.T.: 4.528 min
Delta R.T.: -0.007 min
Response: 96571
Conc: 290.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#9 AR-1221-2

R.T.: 3.850 min
Delta R.T.: -0.002 min
Response: 87077
Conc: 311.99 ng/ml




#10 AR-1221-3

R.T.: 4.603 min
Delta R.T.: -0.007 min
Response: 377440
Conc: 357.55 ng/ml



#10 AR-1221-3


R.T.: 3.923 min
Delta R.T.: -0.002 min
Response: 322760
Conc: 368.51 ng/ml



#11 AR-1232-1


R.T.: 4.603 min
Delta R.T.: 0.000 min
Response: 377440
Conc: 420.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#11 AR-1232-1

R.T.: 3.923 min
Delta R.T.: -0.001 min
Response: 322760
Conc: 432.04 ng/ml




#12 AR-1232-2

R.T.: 5.125 min
Delta R.T.: 0.000 min
Response: 558193
Conc: 1151.89 ng/ml



#12 AR-1232-2


R.T.: 4.639 min
Delta R.T.: -0.001 min
Response: 955865
Conc: 1160.13 ng/ml



#13 AR-1232-3


R.T.: 5.412 min
Delta R.T.: 0.000 min
Response: 1224287
Conc: 1167.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#13 AR-1232-3

R.T.: 4.813 min
Delta R.T.: -0.001 min
Response: 522746
Conc: 1210.24 ng/ml




#14 AR-1232-4

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 615588
Conc: 1158.17 ng/ml



#14 AR-1232-4


R.T.: 4.894 min
Delta R.T.: -0.001 min
Response: 517871
Conc: 1362.09 ng/ml



#15 AR-1232-5


R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 563225
Conc: 1363.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#15 AR-1232-5

R.T.: 5.064 min
Delta R.T.: -0.002 min
Response: 561578
Conc: 1319.07 ng/ml




#16 AR-1242-1

R.T.: 5.390 min
Delta R.T.: -0.006 min
Response: 830689
Conc: 586.13 ng/ml



#16 AR-1242-1


R.T.: 4.621 min
Delta R.T.: -0.002 min
Response: 650638
Conc: 614.80 ng/ml



#17 AR-1242-2


R.T.: 5.412 min
Delta R.T.: -0.007 min
Response: 1224287
Conc: 588.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#17 AR-1242-2

R.T.: 4.639 min
Delta R.T.: -0.002 min
Response: 955865
Conc: 613.54 ng/ml




#18 AR-1242-3

R.T.: 5.473 min
Delta R.T.: -0.006 min
Response: 747902
Conc: 592.42 ng/ml



#18 AR-1242-3


R.T.: 4.813 min
Delta R.T.: -0.002 min
Response: 522746
Conc: 616.52 ng/ml



#19 AR-1242-4


R.T.: 5.571 min
Delta R.T.: -0.006 min
Response: 615588
Conc: 567.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#19 AR-1242-4

R.T.: 4.894 min
Delta R.T.: -0.001 min
Response: 517871
Conc: 620.58 ng/ml




#20 AR-1242-5

R.T.: 6.301 min
Delta R.T.: -0.005 min
Response: 112253
Conc: 86.51 ng/ml



#20 AR-1242-5


R.T.: 5.410 min
Delta R.T.: -0.004 min
Response: 429188
Conc: 428.56 ng/ml



#21 AR-1248-1


R.T.: 5.390 min
Delta R.T.: -0.006 min
Response: 830689
Conc: 713.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#21 AR-1248-1

R.T.: 4.621 min
Delta R.T.: -0.002 min
Response: 650638
Conc: 701.68 ng/ml




#22 AR-1248-2

R.T.: 5.660 min
Delta R.T.: -0.006 min
Response: 563225
Conc: 324.46 ng/ml



#22 AR-1248-2


R.T.: 4.853 min
Delta R.T.: -0.002 min
Response: 434544
Conc: 338.07 ng/ml



#23 AR-1248-3


R.T.: 5.862 min
Delta R.T.: -0.006 min
Response: 650864
Conc: 334.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#23 AR-1248-3

R.T.: 4.894 min
Delta R.T.: -0.002 min
Response: 517871
Conc: 392.45 ng/ml




#24 AR-1248-4

R.T.: 6.262 min
Delta R.T.: -0.007 min
Response: 138729
Conc: 59.19 ng/ml



#24 AR-1248-4


R.T.: 5.064 min
Delta R.T.: -0.003 min
Response: 561578
Conc: 346.98 ng/ml



#25 AR-1248-5


R.T.: 6.301 min
Delta R.T.: -0.005 min
Response: 112253
Conc: 49.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#25 AR-1248-5

R.T.: 5.451 min
Delta R.T.: -0.003 min
Response: 72134
Conc: 46.64 ng/ml




#26 AR-1254-1

R.T.: 6.234 min
Delta R.T.: -0.007 min
Response: 489217
Conc: 251.17 ng/ml



#26 AR-1254-1


R.T.: 5.410 min
Delta R.T.: -0.002 min
Response: 429188
Conc: 201.81 ng/ml



#27 AR-1254-2


R.T.: 6.450 min
Delta R.T.: -0.006 min
Response: 554351
Conc: 177.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#27 AR-1254-2

R.T.: 5.556 min
Delta R.T.: -0.002 min
Response: 384095
Conc: 206.36 ng/ml




#28 AR-1254-3

R.T.: 6.815 min
Delta R.T.: -0.007 min
Response: 317633
Conc: 96.08 ng/ml



#28 AR-1254-3


R.T.: 5.968 min
Delta R.T.: 0.011 min
Response: 789570
Conc: 262.21 ng/ml



#29 AR-1254-4


R.T.: 7.098 min
Delta R.T.: -0.007 min
Response: 251310
Conc: 92.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#29 AR-1254-4

R.T.: 6.182 min
Delta R.T.: -0.003 min
Response: 104027
Conc: 57.79 ng/ml




#30 AR-1254-5

R.T.: 7.514 min
Delta R.T.: -0.007 min
Response: 1926323
Conc: 688.98 ng/ml



#30 AR-1254-5


R.T.: 6.594 min
Delta R.T.: -0.003 min
Response: 1345277
Conc: 496.58 ng/ml



#31 AR-1260-1


R.T.: 6.976 min
Delta R.T.: 0.000 min
Response: 1327455
Conc: 478.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#31 AR-1260-1

R.T.: 6.083 min
Delta R.T.: -0.001 min
Response: 1050111
Conc: 490.64 ng/ml




#32 AR-1260-2

R.T.: 7.232 min
Delta R.T.: 0.000 min
Response: 1722579
Conc: 467.25 ng/ml



#32 AR-1260-2


R.T.: 6.270 min
Delta R.T.: -0.001 min
Response: 1329621
Conc: 484.62 ng/ml



#33 AR-1260-3


R.T.: 7.587 min
Delta R.T.: 0.000 min
Response: 1463901
Conc: 481.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#33 AR-1260-3

R.T.: 6.421 min
Delta R.T.: -0.002 min
Response: 1202262
Conc: 492.43 ng/ml




#34 AR-1260-4

R.T.: 7.813 min
Delta R.T.: 0.001 min
Response: 1446692
Conc: 477.03 ng/ml



#34 AR-1260-4


R.T.: 6.887 min
Delta R.T.: -0.001 min
Response: 1031228
Conc: 495.30 ng/ml



#35 AR-1260-5


R.T.: 8.119 min
Delta R.T.: 0.000 min
Response: 2986496
Conc: 477.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#35 AR-1260-5

R.T.: 7.130 min
Delta R.T.: -0.001 min
Response: 2510138
Conc: 494.89 ng/ml




#36 AR-1262-1

R.T.: 7.587 min
Delta R.T.: -0.006 min
Response: 1463901
Conc: 322.10 ng/ml



#36 AR-1262-1


R.T.: 6.630 min
Delta R.T.: -0.003 min
Response: 1272756
Conc: 340.02 ng/ml



#37 AR-1262-2


R.T.: 8.119 min
Delta R.T.: -0.006 min
Response: 2986496
Conc: 416.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#37 AR-1262-2

R.T.: 7.130 min
Delta R.T.: -0.003 min
Response: 2510138
Conc: 422.45 ng/ml




#38 AR-1262-3

R.T.: 8.426 min
Delta R.T.: 0.013 min
Response: 1867377
Conc: 380.31 ng/ml



#38 AR-1262-3


R.T.: 7.408 min
Delta R.T.: -0.002 min
Response: 620249
Conc: 242.49 ng/ml



#39 AR-1262-4


R.T.: 8.477 min
Delta R.T.: -0.023 min
Response: 1329106
Conc: 326.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#39 AR-1262-4

R.T.: 7.472 min
Delta R.T.: -0.003 min
Response: 1757204
Conc: 398.70 ng/ml




#40 AR-1262-5

R.T.: 9.104 min
Delta R.T.: -0.008 min
Response: 868876
Conc: 344.56 ng/ml



#40 AR-1262-5


R.T.: 7.965 min
Delta R.T.: -0.003 min
Response: 623171
Conc: 341.23 ng/ml



#41 AR-1268-1


R.T.: 8.426 min
Delta R.T.: 0.015 min
Response: 1867377
Conc: 221.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#41 AR-1268-1

R.T.: 7.408 min
Delta R.T.: -0.002 min
Response: 620249
Conc: 90.59 ng/ml




#42 AR-1268-2

R.T.: 8.477 min
Delta R.T.: -0.026 min
Response: 1329106
Conc: 166.46 ng/ml



#42 AR-1268-2


R.T.: 7.472 min
Delta R.T.: -0.004 min
Response: 1757204
Conc: 276.85 ng/ml



#43 AR-1268-3


R.T.: 8.701 min
Delta R.T.: -0.007 min
Response: 56242
Conc: 8.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419




#43 AR-1268-3

R.T.: 7.673 min
Delta R.T.: -0.003 min
Response: 40336
Conc: 7.54 ng/ml




#44 AR-1268-4

R.T.: 9.104 min
Delta R.T.: -0.007 min
Response: 868876
Conc: 337.22 ng/ml



#44 AR-1268-4


R.T.: 7.965 min
Delta R.T.: -0.004 min
Response: 623171
Conc: 311.06 ng/ml



#45 AR-1268-5

R.T.: 9.480 min
Delta R.T.: -0.007 min
Response: 267816
Conc: 13.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062419



#45 AR-1268-5

R.T.: 8.239 min
Delta R.T.: -0.003 min
Response: 170750
Conc: 11.11 ng/ml