

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0062518.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jun 25 12:49:48 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046188.D 1000 =P0046186.D 750 =P0046187.D
 250 =P0046189.D 50 =P0046190.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		6.226	5.295	5.960	6.710	5.033	5.845	E5	11.70
2) SA	Decachlorobiphenyl		4.896	4.102	4.631	5.340	4.085	4.611	E5	11.63
3) L1	AR-1016-1		1.482	1.201	1.326	1.596	1.269	1.375	E4	11.73
4) L1	AR-1016-2		1.983	1.612	1.781	2.155	1.692	1.845	E4	12.03
5) L1	AR-1016-3		1.659	1.330	1.476	1.771	1.349	1.517	E4	12.75
6) L1	AR-1016-4		1.650	1.372	1.548	1.855	1.459	1.577	E4	11.84
7) L1	AR-1016-5		1.387	1.137	1.276	1.568	1.242	1.322	E4	12.40
8) L2	AR-1221-1		7.446					7.446	E3	0.00
9) L2	AR-1221-2		5.422					5.422	E3	0.00
10) L2	AR-1221-3		1.672					1.672	E4	0.00
11) L3	AR-1232-1		6.101					6.101	E3	0.00
12) L3	AR-1232-2		8.678					8.678	E3	0.00
13) L3	AR-1232-3		1.269					1.269	E4	0.00
14) L3	AR-1232-4		8.010					8.010	E3	0.00
15) L3	AR-1232-5		6.425					6.425	E3	0.00
16) L4	AR-1242-1		2.504					2.504	E4	0.00
17) L4	AR-1242-2		1.580					1.580	E4	0.00
18) L4	AR-1242-3		1.295					1.295	E4	0.00
19) L4	AR-1242-4		1.223					1.223	E4	0.00
20) L4	AR-1242-5		1.408					1.408	E4	0.00
21) L5	AR-1248-1		1.959					1.959	E4	0.00
22) L5	AR-1248-2		1.765					1.765	E4	0.00
23) L5	AR-1248-3		2.311					2.311	E4	0.00
24) L5	AR-1248-4		2.295					2.295	E4	0.00
25) L5	AR-1248-5		1.527					1.527	E4	0.00
26) L6	AR-1254-1		3.428					3.428	E4	0.00
27) L6	AR-1254-2		2.084					2.084	E4	0.00
28) L6	AR-1254-3		3.742					3.742	E4	0.00
29) L6	AR-1254-4		2.740					2.740	E4	0.00
30) L6	AR-1254-5		2.013					2.013	E4	0.00
31) L7	AR-1260-1		2.937	2.469	2.774	3.291	2.656	2.825	E4	11.01
32) L7	AR-1260-2		3.473	2.878	3.315	3.868	3.050	3.317	E4	11.59
33) L7	AR-1260-3		4.131	3.468	3.945	4.509	3.493	3.909	E4	11.28
34) L7	AR-1260-4		5.657	4.831	5.443	5.981	4.450	5.272	E4	11.81
35) L7	AR-1260-5		3.432	2.927	3.292	3.671	2.831	3.231	E4	10.84
36) L8	AR-1262-1		2.186					2.186	E4	0.00
37) L8	AR-1262-2		2.618					2.618	E4	0.00
38) L8	AR-1262-3		2.408					2.408	E4	0.00
39) L8	AR-1262-4		3.358					3.358	E4	0.00
40) L8	AR-1262-5		6.246					6.246	E4	0.00
41) L9	AR-1268-1		6.214					6.214	E4	0.00
42) L9	AR-1268-2		5.783					5.783	E4	0.00
43) L9	AR-1268-3		4.768					4.768	E4	0.00
44) L9	AR-1268-4		1.945					1.945	E4	0.00
45) L9	AR-1268-5		1.492					1.492	E5	0.00

Signal #2 Calibration Files

500 =P0046188.D 1000 =P0046186.D 750 =P0046187.D
 250 =P0046189.D 50 =P0046190.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		2.594	2.096	2.417	2.948	2.355	2.482	E5	12.74
2) SA	Decachlorobiphenyl		1.760	1.381	1.596	2.011	1.616	1.673	E5	13.89

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0062518.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jun 25 12:49:48 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046188.D 1000 =P0046186.D 750 =P0046187.D
 250 =P0046189.D 50 =P0046190.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.162	5.714	6.361	8.258	6.368	6.773	E3	14.41
4)	L1	AR-1016-2	5.579	4.421	5.172	6.354	4.915	5.288	E3	13.78
5)	L1	AR-1016-3	6.799	5.408	6.301	7.642	4.790	6.188	E3	18.18
6)	L1	AR-1016-4	7.055	5.646	6.631	8.238	6.320	6.778	E3	14.23
7)	L1	AR-1016-5	6.830	5.508	6.310	7.853	5.221	6.344	E3	16.67
8)	L2	AR-1221-1	3.663					3.663	E3	0.00
9)	L2	AR-1221-2	2.636					2.636	E3	0.00
10)	L2	AR-1221-3	7.904					7.904	E3	0.00
11)	L3	AR-1232-1	2.879					2.879	E3	0.00
12)	L3	AR-1232-2	3.844					3.844	E3	0.00
13)	L3	AR-1232-3	5.679					5.679	E3	0.00
14)	L3	AR-1232-4	4.064					4.064	E3	0.00
15)	L3	AR-1232-5	2.692					2.692	E3	0.00
16)	L4	AR-1242-1	5.434					5.434	E3	0.00
17)	L4	AR-1242-2	1.041					1.041	E4	0.00
18)	L4	AR-1242-3	5.571					5.571	E3	0.00
19)	L4	AR-1242-4	5.418					5.418	E3	0.00
20)	L4	AR-1242-5	5.768					5.768	E3	0.00
21)	L5	AR-1248-1	8.783					8.783	E3	0.00
22)	L5	AR-1248-2	8.081					8.081	E3	0.00
23)	L5	AR-1248-3	1.222					1.222	E4	0.00
24)	L5	AR-1248-4	9.002					9.002	E3	0.00
25)	L5	AR-1248-5	5.920					5.920	E3	0.00
26)	L6	AR-1254-1	1.265					1.265	E4	0.00
27)	L6	AR-1254-2	1.119					1.119	E4	0.00
28)	L6	AR-1254-3	1.785					1.785	E4	0.00
29)	L6	AR-1254-4	1.128					1.128	E4	0.00
30)	L6	AR-1254-5	7.856					7.856	E3	0.00
31)	L7	AR-1260-1	1.270	1.018	1.156	1.476	1.245	1.233	E4	13.59
32)	L7	AR-1260-2	1.521	1.213	1.392	1.771	1.528	1.485	E4	13.79
33)	L7	AR-1260-3	1.640	1.321	1.504	1.879	1.520	1.573	E4	13.08
34)	L7	AR-1260-4	2.457	1.979	2.258	2.784	2.156	2.327	E4	13.26
35)	L7	AR-1260-5	1.793	1.434	1.654	2.031	1.574	1.697	E4	13.41
36)	L8	AR-1262-1	1.141					1.141	E4	0.00
37)	L8	AR-1262-2	1.129					1.129	E4	0.00
38)	L8	AR-1262-3	1.553					1.553	E4	0.00
39)	L8	AR-1262-4	1.369					1.369	E4	0.00
40)	L8	AR-1262-5	2.511					2.511	E4	0.00
41)	L9	AR-1268-1	2.443					2.443	E4	0.00
42)	L9	AR-1268-2	2.422					2.422	E4	0.00
43)	L9	AR-1268-3	1.880					1.880	E4	0.00
44)	L9	AR-1268-4	8.366					8.366	E3	0.00
45)	L9	AR-1268-5	4.999					4.999	E4	0.00

(#) = Out of Range