

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062518\
 Data File : P0046190.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2018 10:34
 Operator : SM/UA
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

ugo
 6/26/2018 10:09:11 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 11:02:10 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 25 11:00:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.378	3.605	2516312	1177474	4.305	4.744
2)	SA Decachlor...	10.047	8.576	2042293	807762	4.430	4.819m
Target Compounds							
3)	L1 AR-1016-1	5.273	4.465	634486	318390	46.157	47.012
4)	L1 AR-1016-2	5.623	4.921	846114	245749	45.870	46.676m
5)	L1 AR-1016-3	5.720	4.962	674456	239495	44.455	38.703
6)	L1 AR-1016-4	6.154	5.133	729624	316012	46.273	46.418
7)	L1 AR-1016-5	6.389	5.482	621034	261031	46.869m	41.144
31)	L7 AR-1260-1	7.136	6.159	1327948	622677	47.002	50.504
32)	L7 AR-1260-2	7.391	6.345	1525239	763791	45.984	51.433
33)	L7 AR-1260-3	7.671	6.672	1746275	759929	44.673	48.315
34)	L7 AR-1260-4	8.286	7.208	2224806	1077931	42.199	46.328
35)	L7 AR-1260-5	8.602	7.553	1415447	787015	43.811	46.369

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062518\
Data File : P0046190.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2018 10:34
Operator : SM/UA
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 11:02:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 25 11:00:58 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.378 min
Delta R.T.: 0.000 min
Response: 2516312
Conc: 4.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 1177474
Conc: 4.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.047 min
Delta R.T.: 0.000 min
Response: 2042293
Conc: 4.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.576 min
Delta R.T.: 0.000 min
Response: 807762
Conc: 4.82 ng/ml m




#3 AR-1016-1

R.T.: 5.273 min
Delta R.T.: 0.000 min
Response: 634486
Conc: 46.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#3 AR-1016-1

R.T.: 4.465 min
Delta R.T.: 0.000 min
Response: 318390
Conc: 47.01 ng/ml




#4 AR-1016-2

R.T.: 5.623 min
Delta R.T.: 0.000 min
Response: 846114
Conc: 45.87 ng/ml



#4 AR-1016-2

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 245749
Conc: 46.68 ng/ml m




#5 AR-1016-3

R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 674456
Conc: 44.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#5 AR-1016-3

R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 239495
Conc: 38.70 ng/ml




#6 AR-1016-4

R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 729624
Conc: 46.27 ng/ml



#6 AR-1016-4

R.T.: 5.133 min
Delta R.T.: 0.000 min
Response: 316012
Conc: 46.42 ng/ml




#7 AR-1016-5

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 621034
Conc: 46.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#7 AR-1016-5

R.T.: 5.482 min
Delta R.T.: 0.000 min
Response: 261031
Conc: 41.14 ng/ml




#31 AR-1260-1

R.T.: 7.136 min
Delta R.T.: 0.000 min
Response: 1327948
Conc: 47.00 ng/ml



#31 AR-1260-1

R.T.: 6.159 min
Delta R.T.: 0.000 min
Response: 622677
Conc: 50.50 ng/ml




#32 AR-1260-2

R.T.: 7.391 min
Delta R.T.: 0.000 min
Response: 1525239
Conc: 45.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#32 AR-1260-2

R.T.: 6.345 min
Delta R.T.: 0.000 min
Response: 763791
Conc: 51.43 ng/ml




#33 AR-1260-3

R.T.: 7.671 min
Delta R.T.: 0.000 min
Response: 1746275
Conc: 44.67 ng/ml



#33 AR-1260-3

R.T.: 6.672 min
Delta R.T.: 0.000 min
Response: 759929
Conc: 48.32 ng/ml




#34 AR-1260-4

R.T.: 8.286 min
Delta R.T.: 0.000 min
Response: 2224806
Conc: 42.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

ugo
6/26/2018 10:09:11 AM




#34 AR-1260-4

R.T.: 7.208 min
Delta R.T.: 0.000 min
Response: 1077931
Conc: 46.33 ng/ml



#35 AR-1260-5

R.T.: 8.602 min
Delta R.T.: 0.000 min
Response: 1415447
Conc: 43.81 ng/ml



#35 AR-1260-5

R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 787015
Conc: 46.37 ng/ml