

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062519\
 Data File : P0057460.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2019 23:27
 Operator : SM/SJ
 Sample : K3500-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B-1(9-10)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 02:02:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.226	3.554	888477	768724	22.618	21.862
2) SA Decachlor...	9.774	8.473	1158952	803755	20.610	21.487

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062519\
Data File : PO057460.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2019 23:27
Operator : SM/SJ
Sample : K3500-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-1(9-10)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 02:02:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.226 min
Delta R.T.: -0.004 min
Response: 888477
Conc: 22.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-1(9-10)




#1 Tetrachloro-m-xylene

R.T.: 3.554 min
Delta R.T.: -0.002 min
Response: 768724
Conc: 21.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.774 min
Delta R.T.: -0.007 min
Response: 1158952
Conc: 20.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.473 min
Delta R.T.: -0.004 min
Response: 803755
Conc: 21.49 ng/ml