

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070819\
 Data File : P0057750.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jul 2019 18:56
 Operator : SM/SJ
 Sample : PB121184BSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB121184BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 01:20:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.212	3.546	772339	690595	19.662	19.640
2)	SA Decachlor...	9.742	8.458	1177133	794699	20.933	21.245
Target Compounds							
3)	L1 AR-1016-1	5.391	4.628	361217	314392	204.371	235.150
4)	L1 AR-1016-2	5.391	4.628	361217	314392	140.170	161.485
5)	L1 AR-1016-3	5.452	4.801	315880	171656	202.636	162.133
6)	L1 AR-1016-4	5.639	4.842	223649	155288	172.464	174.538
7)	L1 AR-1016-5	5.880	5.052	1083	212059	0.792	185.625 #
8)	L2 AR-1221-1	4.510	0.000	35518	0	81.587	N.D. #
9)	L2 AR-1221-2	4.510	3.914	35518	123786	106.885	443.517 #
10)	L2 AR-1221-3	4.585	3.914	158668	123786	150.307	141.333
11)	L3 AR-1232-1	4.585	3.914	158668	123786	176.749	165.698
12)	L3 AR-1232-2	5.105	4.628	242094	314392	499.588	381.576
13)	L3 AR-1232-3	5.391	4.801	361217	171656	344.410	397.411
14)	L3 AR-1232-4	5.639	4.882	223649	199959	420.773	525.929
15)	L3 AR-1232-5	5.639	5.052	223649	212059	541.606	498.096
16)	L4 AR-1242-1	5.391	4.628	361217	314392	254.872	297.073
17)	L4 AR-1242-2	5.391	4.628	361217	314392	173.674	201.798
18)	L4 AR-1242-3	5.452	4.801	315880	171656	250.212	202.450
19)	L4 AR-1242-4	5.639	4.882	223649	199959	206.144	239.619
20)	L4 AR-1242-5	6.240	5.398	39529	180626	30.464	180.363 #
21)	L5 AR-1248-1	5.391	4.628	361217	314392	310.132	339.056
22)	L5 AR-1248-2	5.683	4.842	243615	155288	140.340	120.813
23)	L5 AR-1248-3	5.880	4.882	1083	199959	0.556	151.530 #
24)	L5 AR-1248-4	6.240	5.052	39529	212059	16.865	131.025 #
25)	L5 AR-1248-5	6.240	5.398	39529	180626	17.603	116.792 #
26)	L6 AR-1254-1	6.240	5.398	39529	180626	20.295	84.932 #
27)	L6 AR-1254-2	6.427	5.544	262559	182358	84.241	97.975
28)	L6 AR-1254-3	6.848	5.954	426776	351192	129.089	116.628
29)	L6 AR-1254-4	7.075	6.197	99039	119930	36.493	66.626 #
30)	L6 AR-1254-5	7.522	6.581	126983	671687	45.418	247.937 #
31)	L7 AR-1260-1	6.952	6.069	633756	496615	228.680	232.031
32)	L7 AR-1260-2	7.253	6.257	252634	629604	68.528	229.479 #
33)	L7 AR-1260-3	7.563	6.407	624133	566225	205.433	231.916
34)	L7 AR-1260-4	7.788	6.873	617239	430867	203.526	206.945
35)	L7 AR-1260-5	8.141	7.116	83844	970441	13.407	191.328 #
36)	L8 AR-1262-1	7.620	6.616	324485	531348	71.397	141.949 #
37)	L8 AR-1262-2	8.141	7.116	83844	970441	11.702	163.323 #
38)	L8 AR-1262-3	8.400	7.392	764022	230631	155.602	90.167 #
39)	L8 AR-1262-4	8.448	7.457	484760	705504	119.155	160.074 #
40)	L8 AR-1262-5	9.071	7.950	288302	187941	114.329	102.911
41)	L9 AR-1268-1	8.400	7.392	764022	230631	90.609	33.684 #
42)	L9 AR-1268-2	8.448	7.457	484760	705504	60.713	111.154 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070819\
 Data File : P0057750.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jul 2019 18:56
 Operator : SM/SJ
 Sample : PB121184BSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB121184BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 01:20:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	9.071	7.950	288302	187941	111.893	93.812
45) L9	AR-1268-5	9.444	8.223	101787	61142	5.146	3.979



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070819\
Data File : PO057750.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jul 2019 18:56
Operator : SM/SJ
Sample : PB121184BSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB121184BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 01:20:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.212 min
Delta R.T.: -0.018 min
Response: 772339
Conc: 19.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: -0.009 min
Response: 690595
Conc: 19.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.742 min
Delta R.T.: -0.039 min
Response: 1177133
Conc: 20.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.458 min
Delta R.T.: -0.019 min
Response: 794699
Conc: 21.24 ng/ml



#3 AR-1016-1


R.T.: 5.391 min
Delta R.T.: 0.000 min
Response: 361217
Conc: 204.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#3 AR-1016-1

R.T.: 4.628 min
Delta R.T.: 0.006 min
Response: 314392
Conc: 235.15 ng/ml




#4 AR-1016-2

R.T.: 5.391 min
Delta R.T.: -0.021 min
Response: 361217
Conc: 140.17 ng/ml



#4 AR-1016-2


R.T.: 4.628 min
Delta R.T.: -0.012 min
Response: 314392
Conc: 161.48 ng/ml



#5 AR-1016-3


R.T.: 5.452 min
Delta R.T.: -0.019 min
Response: 315880
Conc: 202.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#5 AR-1016-3

R.T.: 4.801 min
Delta R.T.: -0.012 min
Response: 171656
Conc: 162.13 ng/ml




#6 AR-1016-4

R.T.: 5.639 min
Delta R.T.: 0.067 min
Response: 223649
Conc: 172.46 ng/ml



#6 AR-1016-4


R.T.: 4.842 min
Delta R.T.: -0.013 min
Response: 155288
Conc: 174.54 ng/ml



#7 AR-1016-5


R.T.: 5.880 min
Delta R.T.: 0.020 min
Response: 1083
Conc: 0.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#7 AR-1016-5

R.T.: 5.052 min
Delta R.T.: -0.012 min
Response: 212059
Conc: 185.63 ng/ml




#8 AR-1221-1

R.T.: 4.510 min
Delta R.T.: 0.062 min
Response: 35518
Conc: 81.59 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.766 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.510 min
Delta R.T.: -0.024 min
Response: 35518
Conc: 106.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#9 AR-1221-2

R.T.: 3.914 min
Delta R.T.: 0.063 min
Response: 123786
Conc: 443.52 ng/ml




#10 AR-1221-3

R.T.: 4.585 min
Delta R.T.: -0.025 min
Response: 158668
Conc: 150.31 ng/ml



#10 AR-1221-3


R.T.: 3.914 min
Delta R.T.: -0.012 min
Response: 123786
Conc: 141.33 ng/ml



#11 AR-1232-1


R.T.: 4.585 min
Delta R.T.: -0.019 min
Response: 158668
Conc: 176.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#11 AR-1232-1

R.T.: 3.914 min
Delta R.T.: -0.011 min
Response: 123786
Conc: 165.70 ng/ml




#12 AR-1232-2

R.T.: 5.105 min
Delta R.T.: -0.021 min
Response: 242094
Conc: 499.59 ng/ml



#12 AR-1232-2


R.T.: 4.628 min
Delta R.T.: -0.012 min
Response: 314392
Conc: 381.58 ng/ml



#13 AR-1232-3


R.T.: 5.391 min
Delta R.T.: -0.021 min
Response: 361217
Conc: 344.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#13 AR-1232-3

R.T.: 4.801 min
Delta R.T.: -0.013 min
Response: 171656
Conc: 397.41 ng/ml




#14 AR-1232-4

R.T.: 5.639 min
Delta R.T.: 0.067 min
Response: 223649
Conc: 420.77 ng/ml



#14 AR-1232-4


R.T.: 4.882 min
Delta R.T.: -0.013 min
Response: 199959
Conc: 525.93 ng/ml



#15 AR-1232-5


R.T.: 5.639 min
Delta R.T.: -0.021 min
Response: 223649
Conc: 541.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#15 AR-1232-5

R.T.: 5.052 min
Delta R.T.: -0.014 min
Response: 212059
Conc: 498.10 ng/ml




#16 AR-1242-1

R.T.: 5.391 min
Delta R.T.: -0.005 min
Response: 361217
Conc: 254.87 ng/ml



#16 AR-1242-1


R.T.: 4.628 min
Delta R.T.: 0.005 min
Response: 314392
Conc: 297.07 ng/ml



#17 AR-1242-2


R.T.: 5.391 min
Delta R.T.: -0.027 min
Response: 361217
Conc: 173.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#17 AR-1242-2

R.T.: 4.628 min
Delta R.T.: -0.013 min
Response: 314392
Conc: 201.80 ng/ml




#18 AR-1242-3

R.T.: 5.452 min
Delta R.T.: -0.027 min
Response: 315880
Conc: 250.21 ng/ml



#18 AR-1242-3


R.T.: 4.801 min
Delta R.T.: -0.014 min
Response: 171656
Conc: 202.45 ng/ml



#19 AR-1242-4


R.T.: 5.639 min
Delta R.T.: 0.061 min
Response: 223649
Conc: 206.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#19 AR-1242-4

R.T.: 4.882 min
Delta R.T.: -0.014 min
Response: 199959
Conc: 239.62 ng/ml




#20 AR-1242-5

R.T.: 6.240 min
Delta R.T.: -0.066 min
Response: 39529
Conc: 30.46 ng/ml



#20 AR-1242-5


R.T.: 5.398 min
Delta R.T.: -0.016 min
Response: 180626
Conc: 180.36 ng/ml



#21 AR-1248-1


R.T.: 5.391 min
Delta R.T.: -0.005 min
Response: 361217
Conc: 310.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#21 AR-1248-1

R.T.: 4.628 min
Delta R.T.: 0.005 min
Response: 314392
Conc: 339.06 ng/ml




#22 AR-1248-2

R.T.: 5.683 min
Delta R.T.: 0.017 min
Response: 243615
Conc: 140.34 ng/ml



#22 AR-1248-2


R.T.: 4.842 min
Delta R.T.: -0.014 min
Response: 155288
Conc: 120.81 ng/ml



#23 AR-1248-3


R.T.: 5.880 min
Delta R.T.: 0.012 min
Response: 1083
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#23 AR-1248-3

R.T.: 4.882 min
Delta R.T.: -0.014 min
Response: 199959
Conc: 151.53 ng/ml




#24 AR-1248-4

R.T.: 6.240 min
Delta R.T.: -0.029 min
Response: 39529
Conc: 16.87 ng/ml



#24 AR-1248-4


R.T.: 5.052 min
Delta R.T.: -0.014 min
Response: 212059
Conc: 131.02 ng/ml



#25 AR-1248-5


R.T.: 6.240 min
Delta R.T.: -0.066 min
Response: 39529
Conc: 17.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#25 AR-1248-5

R.T.: 5.398 min
Delta R.T.: -0.056 min
Response: 180626
Conc: 116.79 ng/ml




#26 AR-1254-1

R.T.: 6.240 min
Delta R.T.: 0.000 min
Response: 39529
Conc: 20.29 ng/ml



#26 AR-1254-1


R.T.: 5.398 min
Delta R.T.: -0.014 min
Response: 180626
Conc: 84.93 ng/ml



#27 AR-1254-2


R.T.: 6.427 min
Delta R.T.: -0.030 min
Response: 262559
Conc: 84.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#27 AR-1254-2

R.T.: 5.544 min
Delta R.T.: -0.014 min
Response: 182358
Conc: 97.98 ng/ml




#28 AR-1254-3

R.T.: 6.848 min
Delta R.T.: 0.025 min
Response: 426776
Conc: 129.09 ng/ml



#28 AR-1254-3


R.T.: 5.954 min
Delta R.T.: -0.003 min
Response: 351192
Conc: 116.63 ng/ml



#29 AR-1254-4


R.T.: 7.075 min
Delta R.T.: -0.031 min
Response: 99039
Conc: 36.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#29 AR-1254-4

R.T.: 6.197 min
Delta R.T.: 0.012 min
Response: 119930
Conc: 66.63 ng/ml




#30 AR-1254-5

R.T.: 7.522 min
Delta R.T.: 0.000 min
Response: 126983
Conc: 45.42 ng/ml



#30 AR-1254-5


R.T.: 6.581 min
Delta R.T.: -0.016 min
Response: 671687
Conc: 247.94 ng/ml



#31 AR-1260-1


R.T.: 6.952 min
Delta R.T.: -0.024 min
Response: 633756
Conc: 228.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#31 AR-1260-1

R.T.: 6.069 min
Delta R.T.: -0.015 min
Response: 496615
Conc: 232.03 ng/ml




#32 AR-1260-2

R.T.: 7.253 min
Delta R.T.: 0.021 min
Response: 252634
Conc: 68.53 ng/ml



#32 AR-1260-2


R.T.: 6.257 min
Delta R.T.: -0.014 min
Response: 629604
Conc: 229.48 ng/ml



#33 AR-1260-3


R.T.: 7.563 min
Delta R.T.: -0.024 min
Response: 624133
Conc: 205.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#33 AR-1260-3

R.T.: 6.407 min
Delta R.T.: -0.015 min
Response: 566225
Conc: 231.92 ng/ml




#34 AR-1260-4

R.T.: 7.788 min
Delta R.T.: -0.023 min
Response: 617239
Conc: 203.53 ng/ml



#34 AR-1260-4


R.T.: 6.873 min
Delta R.T.: -0.015 min
Response: 430867
Conc: 206.94 ng/ml



#35 AR-1260-5


R.T.: 8.141 min
Delta R.T.: 0.021 min
Response: 83844
Conc: 13.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#35 AR-1260-5

R.T.: 7.116 min
Delta R.T.: -0.015 min
Response: 970441
Conc: 191.33 ng/ml




#36 AR-1262-1

R.T.: 7.620 min
Delta R.T.: 0.026 min
Response: 324485
Conc: 71.40 ng/ml



#36 AR-1262-1


R.T.: 6.616 min
Delta R.T.: -0.016 min
Response: 531348
Conc: 141.95 ng/ml



#37 AR-1262-2


R.T.: 8.141 min
Delta R.T.: 0.015 min
Response: 83844
Conc: 11.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#37 AR-1262-2

R.T.: 7.116 min
Delta R.T.: -0.016 min
Response: 970441
Conc: 163.32 ng/ml




#38 AR-1262-3

R.T.: 8.400 min
Delta R.T.: -0.013 min
Response: 764022
Conc: 155.60 ng/ml



#38 AR-1262-3


R.T.: 7.392 min
Delta R.T.: -0.017 min
Response: 230631
Conc: 90.17 ng/ml



#39 AR-1262-4


R.T.: 8.448 min
Delta R.T.: -0.052 min
Response: 484760
Conc: 119.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#39 AR-1262-4

R.T.: 7.457 min
Delta R.T.: -0.018 min
Response: 705504
Conc: 160.07 ng/ml




#40 AR-1262-5

R.T.: 9.071 min
Delta R.T.: -0.041 min
Response: 288302
Conc: 114.33 ng/ml



#40 AR-1262-5


R.T.: 7.950 min
Delta R.T.: -0.018 min
Response: 187941
Conc: 102.91 ng/ml



#41 AR-1268-1


R.T.: 8.400 min
Delta R.T.: -0.012 min
Response: 764022
Conc: 90.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#41 AR-1268-1

R.T.: 7.392 min
Delta R.T.: -0.018 min
Response: 230631
Conc: 33.68 ng/ml




#42 AR-1268-2

R.T.: 8.448 min
Delta R.T.: -0.055 min
Response: 484760
Conc: 60.71 ng/ml



#42 AR-1268-2


R.T.: 7.457 min
Delta R.T.: -0.019 min
Response: 705504
Conc: 111.15 ng/ml



#44 AR-1268-4


R.T.: 9.071 min
Delta R.T.: -0.040 min
Response: 288302
Conc: 111.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB121184BSD




#44 AR-1268-4

R.T.: 7.950 min
Delta R.T.: -0.019 min
Response: 187941
Conc: 93.81 ng/ml



#45 AR-1268-5

R.T.: 9.444 min
Delta R.T.: -0.043 min
Response: 101787
Conc: 5.15 ng/ml



#45 AR-1268-5

R.T.: 8.223 min
Delta R.T.: -0.020 min
Response: 61142
Conc: 3.98 ng/ml