

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0070918\
Data File : P0046482.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Jul 2018 14:33
Operator : SM/UA
Sample : J3872-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I-2D5(12-18)

Manual Integrations
APPROVED

Sohil
7/11/2018 11:31:55 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 06:13:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.213	3.542	5434461	5889690	22.740m	27.728
2)	SA Decachlor...	9.749	8.475	3178453	2910508	16.676m	15.004

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0070918\
Data File : P0046482.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Jul 2018 14:33
Operator : SM/UA
Sample : J3872-01
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
I-2D5(12-18)


Manual Integrations
APPROVED

Sohil
7/11/2018 11:31:55 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 06:13:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.213 min
Delta R.T.: 0.001 min
Response: 5434461
Conc: 22.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
I-2D5(12-18)


Manual Integrations
APPROVED

Sohil
7/11/2018 11:31:55 AM




#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: 0.003 min
Response: 5889690
Conc: 27.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.749 min
Delta R.T.: 0.001 min
Response: 3178453
Conc: 16.68 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.475 min
Delta R.T.: 0.002 min
Response: 2910508
Conc: 15.00 ng/ml