

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071319\
 Data File : P0057906.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jul 2019 19:00
 Operator : SM/AJ
 Sample : K3768-17MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-3MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 01:21:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.204	3.542	710272	640812	18.082	18.224
2)	SA Decachlor...	9.729	8.451	397060	636547	7.061	17.017 #
Target Compounds							
3)	L1 AR-1016-1	5.384	4.622	114695	240672	64.893	180.012 #
4)	L1 AR-1016-2	5.384	4.622	114695	240672	44.508	123.619 #
5)	L1 AR-1016-3	5.444	4.796	321442	100816	206.204	95.223 #
6)	L1 AR-1016-4	5.542	4.837	8581	155524	6.617	174.804 #
7)	L1 AR-1016-5	5.872	5.046	-17785	143937	N.D.	125.995 #
8)	L2 AR-1221-1	4.502	0.000	38056	0	87.416	N.D. #
9)	L2 AR-1221-2	4.502	3.909	38056	121423	114.522	435.052 #
10)	L2 AR-1221-3	4.577	3.909	159848	121423	151.425	138.636
11)	L3 AR-1232-1	4.577	3.909	159848	121423	178.063	162.535
12)	L3 AR-1232-2	5.097	4.622	250624	240672	517.192	292.103 #
13)	L3 AR-1232-3	5.384	4.796	114695	100816	109.359	233.404 #
14)	L3 AR-1232-4	5.542	4.907	8581	65233	16.144	171.575 #
15)	L3 AR-1232-5	5.675	5.046	253222	143937	613.221	338.088 #
16)	L4 AR-1242-1	5.384	4.622	114695	240672	80.928	227.414 #
17)	L4 AR-1242-2	5.444	4.622	321442	240672	154.551	154.480
18)	L4 AR-1242-3	5.444	4.796	321442	100816	254.618	118.901 #
19)	L4 AR-1242-4	5.542	4.907	8581	65233	7.909	78.172 #
20)	L4 AR-1242-5	6.270	5.393	39167	178295	30.185	178.035 #
21)	L5 AR-1248-1	5.384	4.622	114695	240672	98.475	259.553 #
22)	L5 AR-1248-2	5.675	4.837	253222	155524	145.874	120.997
23)	L5 AR-1248-3	5.872	4.907	-17785	65233	N.D.	49.434 #
24)	L5 AR-1248-4	6.270	5.085	39167	60101	16.711	37.135 #
25)	L5 AR-1248-5	6.270	5.393	39167	178295	17.442	115.284 #
26)	L6 AR-1254-1	6.232	5.393	48736	178295	25.022	83.836 #
27)	L6 AR-1254-2	6.419	5.538	257014	177557	82.462	95.396
28)	L6 AR-1254-3	6.838	5.949	413365	334341	125.032	111.032
29)	L6 AR-1254-4	7.130	6.191	126441	112891	46.590	62.715 #
30)	L6 AR-1254-5	7.513	6.610	121174	466115	43.340	172.055 #
31)	L7 AR-1260-1	6.944	6.064	587499	467449	211.989	218.404
32)	L7 AR-1260-2	7.245	6.252	223797	570787	60.706	208.041 #
33)	L7 AR-1260-3	7.612	6.402	275584	518198	90.708	212.245 #
34)	L7 AR-1260-4	7.780	6.907	521361	46595	171.911	22.379 #
35)	L7 AR-1260-5	8.131	7.111	69680	827099	11.142	163.067 #
36)	L8 AR-1262-1	7.612	6.610	275584	466115	60.637	124.522 #
37)	L8 AR-1262-2	8.131	7.111	69680	827099	9.725	139.199 #
38)	L8 AR-1262-3	8.391	7.387	615578	195455	125.370	76.414 #
39)	L8 AR-1262-4	8.439	7.452	374125	595441	91.961	135.102 #
40)	L8 AR-1262-5	9.061	7.945	226944	162317	89.997	88.880
41)	L9 AR-1268-1	8.391	7.387	615578	195455	73.004	28.546 #
42)	L9 AR-1268-2	8.439	7.452	374125	595441	46.857	93.813 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071319\
 Data File : P0057906.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jul 2019 19:00
 Operator : SM/AJ
 Sample : K3768-17MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-3MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 01:21:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	9.061	7.945	226944	162317	88.079	81.021
45) L9	AR-1268-5	9.431	8.218	77619	52361	3.924	3.408



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071319\
Data File : PO057906.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jul 2019 19:00
Operator : SM/AJ
Sample : K3768-17MS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-3MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 01:21:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.204 min
Delta R.T.: -0.026 min
Response: 710272
Conc: 18.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: -0.013 min
Response: 640812
Conc: 18.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.729 min
Delta R.T.: -0.052 min
Response: 397060
Conc: 7.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.451 min
Delta R.T.: -0.025 min
Response: 636547
Conc: 17.02 ng/ml



#3 AR-1016-1


R.T.: 5.384 min
Delta R.T.: -0.007 min
Response: 114695
Conc: 64.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#3 AR-1016-1

R.T.: 4.622 min
Delta R.T.: 0.000 min
Response: 240672
Conc: 180.01 ng/ml




#4 AR-1016-2

R.T.: 5.384 min
Delta R.T.: -0.028 min
Response: 114695
Conc: 44.51 ng/ml



#4 AR-1016-2


R.T.: 4.622 min
Delta R.T.: -0.018 min
Response: 240672
Conc: 123.62 ng/ml



#5 AR-1016-3


R.T.: 5.444 min
Delta R.T.: -0.028 min
Response: 321442
Conc: 206.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#5 AR-1016-3

R.T.: 4.796 min
Delta R.T.: -0.017 min
Response: 100816
Conc: 95.22 ng/ml




#6 AR-1016-4

R.T.: 5.542 min
Delta R.T.: -0.029 min
Response: 8581
Conc: 6.62 ng/ml



#6 AR-1016-4


R.T.: 4.837 min
Delta R.T.: -0.018 min
Response: 155524
Conc: 174.80 ng/ml



#7 AR-1016-5


R.T.: 5.872 min
Delta R.T.: 0.011 min
Response: -17785
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#7 AR-1016-5

R.T.: 5.046 min
Delta R.T.: -0.018 min
Response: 143937
Conc: 125.99 ng/ml




#8 AR-1221-1

R.T.: 4.502 min
Delta R.T.: 0.053 min
Response: 38056
Conc: 87.42 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.766 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.502 min
Delta R.T.: -0.033 min
Response: 38056
Conc: 114.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#9 AR-1221-2

R.T.: 3.909 min
Delta R.T.: 0.058 min
Response: 121423
Conc: 435.05 ng/ml




#10 AR-1221-3

R.T.: 4.577 min
Delta R.T.: -0.033 min
Response: 159848
Conc: 151.43 ng/ml



#10 AR-1221-3


R.T.: 3.909 min
Delta R.T.: -0.016 min
Response: 121423
Conc: 138.64 ng/ml



#11 AR-1232-1


R.T.: 4.577 min
Delta R.T.: -0.027 min
Response: 159848
Conc: 178.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#11 AR-1232-1

R.T.: 3.909 min
Delta R.T.: -0.015 min
Response: 121423
Conc: 162.54 ng/ml




#12 AR-1232-2

R.T.: 5.097 min
Delta R.T.: -0.028 min
Response: 250624
Conc: 517.19 ng/ml



#12 AR-1232-2


R.T.: 4.622 min
Delta R.T.: -0.018 min
Response: 240672
Conc: 292.10 ng/ml



#13 AR-1232-3


R.T.: 5.384 min
Delta R.T.: -0.029 min
Response: 114695
Conc: 109.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#13 AR-1232-3

R.T.: 4.796 min
Delta R.T.: -0.018 min
Response: 100816
Conc: 233.40 ng/ml




#14 AR-1232-4

R.T.: 5.542 min
Delta R.T.: -0.029 min
Response: 8581
Conc: 16.14 ng/ml



#14 AR-1232-4


R.T.: 4.907 min
Delta R.T.: 0.012 min
Response: 65233
Conc: 171.58 ng/ml



#15 AR-1232-5


R.T.: 5.675 min
Delta R.T.: 0.015 min
Response: 253222
Conc: 613.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#15 AR-1232-5

R.T.: 5.046 min
Delta R.T.: -0.019 min
Response: 143937
Conc: 338.09 ng/ml




#16 AR-1242-1

R.T.: 5.384 min
Delta R.T.: -0.013 min
Response: 114695
Conc: 80.93 ng/ml



#16 AR-1242-1


R.T.: 4.622 min
Delta R.T.: 0.000 min
Response: 240672
Conc: 227.41 ng/ml



#17 AR-1242-2


R.T.: 5.444 min
Delta R.T.: 0.026 min
Response: 321442
Conc: 154.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#17 AR-1242-2

R.T.: 4.622 min
Delta R.T.: -0.019 min
Response: 240672
Conc: 154.48 ng/ml




#18 AR-1242-3

R.T.: 5.444 min
Delta R.T.: -0.035 min
Response: 321442
Conc: 254.62 ng/ml



#18 AR-1242-3


R.T.: 4.796 min
Delta R.T.: -0.019 min
Response: 100816
Conc: 118.90 ng/ml



#19 AR-1242-4


R.T.: 5.542 min
Delta R.T.: -0.035 min
Response: 8581
Conc: 7.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#19 AR-1242-4

R.T.: 4.907 min
Delta R.T.: 0.012 min
Response: 65233
Conc: 78.17 ng/ml




#20 AR-1242-5

R.T.: 6.270 min
Delta R.T.: -0.036 min
Response: 39167
Conc: 30.18 ng/ml



#20 AR-1242-5


R.T.: 5.393 min
Delta R.T.: -0.021 min
Response: 178295
Conc: 178.03 ng/ml



#21 AR-1248-1


R.T.: 5.384 min
Delta R.T.: -0.013 min
Response: 114695
Conc: 98.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#21 AR-1248-1

R.T.: 4.622 min
Delta R.T.: 0.000 min
Response: 240672
Conc: 259.55 ng/ml




#22 AR-1248-2

R.T.: 5.675 min
Delta R.T.: 0.009 min
Response: 253222
Conc: 145.87 ng/ml



#22 AR-1248-2


R.T.: 4.837 min
Delta R.T.: -0.019 min
Response: 155524
Conc: 121.00 ng/ml



#23 AR-1248-3


R.T.: 5.872 min
Delta R.T.: 0.004 min
Response: -17785
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#23 AR-1248-3

R.T.: 4.907 min
Delta R.T.: 0.011 min
Response: 65233
Conc: 49.43 ng/ml




#24 AR-1248-4

R.T.: 6.270 min
Delta R.T.: 0.001 min
Response: 39167
Conc: 16.71 ng/ml



#24 AR-1248-4


R.T.: 5.085 min
Delta R.T.: 0.019 min
Response: 60101
Conc: 37.13 ng/ml



#25 AR-1248-5


R.T.: 6.270 min
Delta R.T.: -0.036 min
Response: 39167
Conc: 17.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#25 AR-1248-5

R.T.: 5.393 min
Delta R.T.: -0.061 min
Response: 178295
Conc: 115.28 ng/ml




#26 AR-1254-1

R.T.: 6.232 min
Delta R.T.: -0.008 min
Response: 48736
Conc: 25.02 ng/ml



#26 AR-1254-1


R.T.: 5.393 min
Delta R.T.: -0.019 min
Response: 178295
Conc: 83.84 ng/ml



#27 AR-1254-2


R.T.: 6.419 min
Delta R.T.: -0.038 min
Response: 257014
Conc: 82.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#27 AR-1254-2

R.T.: 5.538 min
Delta R.T.: -0.020 min
Response: 177557
Conc: 95.40 ng/ml




#28 AR-1254-3

R.T.: 6.838 min
Delta R.T.: 0.016 min
Response: 413365
Conc: 125.03 ng/ml



#28 AR-1254-3


R.T.: 5.949 min
Delta R.T.: -0.008 min
Response: 334341
Conc: 111.03 ng/ml



#29 AR-1254-4


R.T.: 7.130 min
Delta R.T.: 0.024 min
Response: 126441
Conc: 46.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#29 AR-1254-4

R.T.: 6.191 min
Delta R.T.: 0.006 min
Response: 112891
Conc: 62.72 ng/ml




#30 AR-1254-5

R.T.: 7.513 min
Delta R.T.: -0.008 min
Response: 121174
Conc: 43.34 ng/ml



#30 AR-1254-5


R.T.: 6.610 min
Delta R.T.: 0.013 min
Response: 466115
Conc: 172.05 ng/ml



#31 AR-1260-1


R.T.: 6.944 min
Delta R.T.: -0.031 min
Response: 587499
Conc: 211.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#31 AR-1260-1

R.T.: 6.064 min
Delta R.T.: -0.020 min
Response: 467449
Conc: 218.40 ng/ml




#32 AR-1260-2

R.T.: 7.245 min
Delta R.T.: 0.013 min
Response: 223797
Conc: 60.71 ng/ml



#32 AR-1260-2


R.T.: 6.252 min
Delta R.T.: -0.020 min
Response: 570787
Conc: 208.04 ng/ml



#33 AR-1260-3


R.T.: 7.612 min
Delta R.T.: 0.026 min
Response: 275584
Conc: 90.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#33 AR-1260-3

R.T.: 6.402 min
Delta R.T.: -0.021 min
Response: 518198
Conc: 212.25 ng/ml




#34 AR-1260-4

R.T.: 7.780 min
Delta R.T.: -0.032 min
Response: 521361
Conc: 171.91 ng/ml



#34 AR-1260-4


R.T.: 6.907 min
Delta R.T.: 0.019 min
Response: 46595
Conc: 22.38 ng/ml



#35 AR-1260-5


R.T.: 8.131 min
Delta R.T.: 0.012 min
Response: 69680
Conc: 11.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#35 AR-1260-5

R.T.: 7.111 min
Delta R.T.: -0.020 min
Response: 827099
Conc: 163.07 ng/ml




#36 AR-1262-1

R.T.: 7.612 min
Delta R.T.: 0.019 min
Response: 275584
Conc: 60.64 ng/ml



#36 AR-1262-1


R.T.: 6.610 min
Delta R.T.: -0.022 min
Response: 466115
Conc: 124.52 ng/ml



#37 AR-1262-2


R.T.: 8.131 min
Delta R.T.: 0.006 min
Response: 69680
Conc: 9.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#37 AR-1262-2

R.T.: 7.111 min
Delta R.T.: -0.021 min
Response: 827099
Conc: 139.20 ng/ml




#38 AR-1262-3

R.T.: 8.391 min
Delta R.T.: -0.022 min
Response: 615578
Conc: 125.37 ng/ml



#38 AR-1262-3


R.T.: 7.387 min
Delta R.T.: -0.023 min
Response: 195455
Conc: 76.41 ng/ml



#39 AR-1262-4


R.T.: 8.439 min
Delta R.T.: -0.061 min
Response: 374125
Conc: 91.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#39 AR-1262-4

R.T.: 7.452 min
Delta R.T.: -0.024 min
Response: 595441
Conc: 135.10 ng/ml




#40 AR-1262-5

R.T.: 9.061 min
Delta R.T.: -0.051 min
Response: 226944
Conc: 90.00 ng/ml



#40 AR-1262-5


R.T.: 7.945 min
Delta R.T.: -0.023 min
Response: 162317
Conc: 88.88 ng/ml



#41 AR-1268-1


R.T.: 8.391 min
Delta R.T.: -0.021 min
Response: 615578
Conc: 73.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#41 AR-1268-1

R.T.: 7.387 min
Delta R.T.: -0.024 min
Response: 195455
Conc: 28.55 ng/ml




#42 AR-1268-2

R.T.: 8.439 min
Delta R.T.: -0.063 min
Response: 374125
Conc: 46.86 ng/ml



#42 AR-1268-2


R.T.: 7.452 min
Delta R.T.: -0.024 min
Response: 595441
Conc: 93.81 ng/ml



#44 AR-1268-4


R.T.: 9.061 min
Delta R.T.: -0.050 min
Response: 226944
Conc: 88.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3MS




#44 AR-1268-4

R.T.: 7.945 min
Delta R.T.: -0.023 min
Response: 162317
Conc: 81.02 ng/ml



#45 AR-1268-5

R.T.: 9.431 min
Delta R.T.: -0.056 min
Response: 77619
Conc: 3.92 ng/ml



#45 AR-1268-5

R.T.: 8.218 min
Delta R.T.: -0.025 min
Response: 52361
Conc: 3.41 ng/ml