

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071319\
 Data File : P0057895.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jul 2019 14:48
 Operator : SM/AJ
 Sample : K3765-02DL 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S9-04-SDL

Manual Integrations
 APPROVED

mohammad
 7/15/2019 3:28:00 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 15:34:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.205	3.543	500800	431496	12.749	12.271
2)	SA Decachlor...	9.731	8.452	546507	393611	9.719m	10.522
Target Compounds							
26)	L6 AR-1254-1	6.208	5.393	1013164	883145	520.170m	415.263m
27)	L6 AR-1254-2	6.421	5.539	1627957	883058	522.322m	474.442
28)	L6 AR-1254-3	6.786	5.937	2063754	1672832	624.232m	555.533
29)	L6 AR-1254-4	7.069	6.164	1747388	1192989	643.865m	662.751m
30)	L6 AR-1254-5	7.484	6.575	2220384	2007989	794.161m	741.200m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071319\
Data File : PO057895.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jul 2019 14:48
Operator : SM/AJ
Sample : K3765-02DL 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S9-04-SDL

Manual Integrations
APPROVED

mohammad
7/15/2019 3:28:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 15:34:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

