

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071619\
 Data File : P0057991.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Jul 2019 11:37
 Operator : SM/AJ
 Sample : K3825-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S12-09-B

Manual Integrations
 APPROVED

Ankita
 7/18/2019 4:54:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 12:49:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 24 14:02:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.199	3.539	756213	762964	19.251	21.698
2)	SA Decachlor...	9.717	8.445	1203557	762647	21.403m	20.388

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071619\
Data File : PO057991.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Jul 2019 11:37
Operator : SM/AJ
Sample : K3825-02
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
S12-09-B


Manual Integrations
APPROVED

Ankita
7/18/2019 4:54:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 12:49:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.199 min
Delta R.T.: -0.031 min
Response: 756213
Conc: 19.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
S12-09-B


Manual Integrations
APPROVED

Ankita
7/18/2019 4:54:16 PM




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: -0.016 min
Response: 762964
Conc: 21.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.717 min
Delta R.T.: -0.064 min
Response: 1203557
Conc: 21.40 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.445 min
Delta R.T.: -0.031 min
Response: 762647
Conc: 20.39 ng/ml