

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071818\
Data File : P0046866.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jul 2018 13:22
Operator : SM/SJ
Sample : J4011-03RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
20180413-CORE-3RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 23:19:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.201	3.525	5000533	6414344	20.924	30.199 #
2) SA Decachlor...	9.727	8.450	2059273	1816994	10.804	9.367

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071818\
Data File : P0046866.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jul 2018 13:22
Operator : SM/SJ
Sample : J4011-03RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
20180413-CORE-3RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 23:19:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.201 min
Delta R.T.: -0.012 min
Response: 5000533
Conc: 20.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
20180413-CORE-3RE




#1 Tetrachloro-m-xylene

R.T.: 3.525 min
Delta R.T.: -0.014 min
Response: 6414344
Conc: 30.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.727 min
Delta R.T.: -0.021 min
Response: 2059273
Conc: 10.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.450 min
Delta R.T.: -0.023 min
Response: 1816994
Conc: 9.37 ng/ml