

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0071818\
Data File : P0046867.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jul 2018 13:38
Operator : SM/SJ
Sample : J4011-04RE
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
20180415-COMPOSITE-1RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 23:19:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.201	3.525	5484548	6146099	22.949	28.936 #
2)	SA Decachlor...	9.727	8.450	2084594	1979970	10.937	10.207

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071818\
Data File : P0046867.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jul 2018 13:38
Operator : SM/SJ
Sample : J4011-04RE
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
20180415-COMPOSITE-1RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 23:19:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.201 min
Delta R.T.: -0.011 min
Response: 5484548
Conc: 22.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
20180415-COMPOSITE-1RE




#1 Tetrachloro-m-xylene

R.T.: 3.525 min
Delta R.T.: -0.014 min
Response: 6146099
Conc: 28.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.727 min
Delta R.T.: -0.021 min
Response: 2084594
Conc: 10.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.450 min
Delta R.T.: -0.023 min
Response: 1979970
Conc: 10.21 ng/ml