

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072419\
Data File : P0058342.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jul 2019 5:03
Operator : SM/AJ
Sample : K3929-06
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
30321

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 06:39:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 17 11:13:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.190	3.532	383156	331644	11.307	10.340
2)	SA Decachlor...	9.692	8.427	291973	207575	5.707	6.056

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072419\
Data File : P0058342.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jul 2019 5:03
Operator : SM/AJ
Sample : K3929-06
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
30321

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 06:39:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 17 11:13:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.190 min
Delta R.T.: -0.010 min
Response: 383156
Conc: 11.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
30321




#1 Tetrachloro-m-xylene

R.T.: 3.532 min
Delta R.T.: -0.008 min
Response: 331644
Conc: 10.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.692 min
Delta R.T.: -0.026 min
Response: 291973
Conc: 5.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.427 min
Delta R.T.: -0.018 min
Response: 207575
Conc: 6.06 ng/ml