

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072519\
 Data File : P0058348.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jul 2019 10:11
 Operator : SM/AJ
 Sample : K3964-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 S11-01-S

Manual Integrations
 APPROVED

Ankita
 7/25/2019 1:59:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 10:47:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0071719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 17 11:13:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.196	3.530	686255	618822	20.251	19.294
2) SA Decachlor...	9.702	8.429	865886	583980	16.925	17.038
Target Compounds						
41) L9 AR-1268-1	8.352	7.369	341154	245706	33.177	31.473
42) L9 AR-1268-2	8.439	7.435	196283	171737	21.430	24.103
43) L9 AR-1268-3	8.641	7.632	390056	301943	48.840m	50.127m
44) L9 AR-1268-4	9.041	7.925	41090	30009	13.867m	14.346m
45) L9 AR-1268-5	9.407	8.198	897172	613028	41.187	36.761m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072519\
Data File : P0058348.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jul 2019 10:11
Operator : SM/AJ
Sample : K3964-01
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
S11-01-S


Manual Integrations
APPROVED

Ankita
7/25/2019 1:59:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 10:47:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 17 11:13:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.196 min
Delta R.T.: -0.004 min
Response: 686255
Conc: 20.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
S11-01-S


Manual Integrations
APPROVED

Ankita
7/25/2019 1:59:14 PM




#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.009 min
Response: 618822
Conc: 19.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.702 min
Delta R.T.: -0.016 min
Response: 865886
Conc: 16.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.429 min
Delta R.T.: -0.016 min
Response: 583980
Conc: 17.04 ng/ml




#41 AR-1268-1

R.T.: 8.352 min
Delta R.T.: -0.009 min
Response: 341154
Conc: 33.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
S11-01-S


Manual Integrations
APPROVED

Ankita
7/25/2019 1:59:14 PM




#41 AR-1268-1

R.T.: 7.369 min
Delta R.T.: -0.012 min
Response: 245706
Conc: 31.47 ng/ml




#42 AR-1268-2

R.T.: 8.439 min
Delta R.T.: -0.010 min
Response: 196283
Conc: 21.43 ng/ml



#42 AR-1268-2

R.T.: 7.435 min
Delta R.T.: -0.013 min
Response: 171737
Conc: 24.10 ng/ml




#43 AR-1268-3

R.T.: 8.641 min
Delta R.T.: -0.012 min
Response: 390056
Conc: 48.84 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S11-01-S


Manual Integrations
APPROVED

Ankita
7/25/2019 1:59:14 PM




#43 AR-1268-3

R.T.: 7.632 min
Delta R.T.: -0.013 min
Response: 301943
Conc: 50.13 ng/ml m




#44 AR-1268-4

R.T.: 9.041 min
Delta R.T.: -0.009 min
Response: 41090
Conc: 13.87 ng/ml m



#44 AR-1268-4

R.T.: 7.925 min
Delta R.T.: -0.014 min
Response: 30009
Conc: 14.35 ng/ml m




#45 AR-1268-5

R.T.: 9.407 min
Delta R.T.: -0.013 min
Response: 897172
Conc: 41.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
S11-01-S

Manual Integrations
APPROVED

Ankita
7/25/2019 1:59:14 PM



#45 AR-1268-5

R.T.: 8.198 min
Delta R.T.: -0.013 min
Response: 613028
Conc: 36.76 ng/ml m