

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081919\
 Data File : P0059463.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Aug 2019 18:19
 Operator : SM/AJ
 Sample : K3591-09
 Misc : AR1660 MDL 25PPB
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MDL-MDL-WATER-03-QT3-2019

Manual Integrations
 APPROVED

Ankita
 8/21/2019 12:58:08 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 01:05:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.152	3.506	485033	448835	13.853	12.469
2)	SA Decachlor...	9.643	8.380	730844	526640	13.738	13.061
Target Compounds							
3)	L1 AR-1016-1	5.305	4.560	40685	26954	25.367m	19.938
4)	L1 AR-1016-2	5.325	4.578	54206	34328	22.657m	17.138
5)	L1 AR-1016-3	5.387	4.746	36821	11785	24.720	10.893m#
6)	L1 AR-1016-4	5.485	4.790	32918	17160	27.211	19.004m#
7)	L1 AR-1016-5	5.774	5.000	35698	18777	27.799	16.301 #
31)	L7 AR-1260-1	6.885	6.010	79073	43638	29.332m	19.149m#
32)	L7 AR-1260-2	7.139	6.198	97322	108819	26.229m	37.141m#
33)	L7 AR-1260-3	7.492	6.346	96697	51909	29.514m	19.595m#
34)	L7 AR-1260-4	7.716	6.810	110358	83404	35.932m	36.862m
35)	L7 AR-1260-5	8.014	7.052	209952	236497	29.650m	42.973m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081919\
Data File : PO059463.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Aug 2019 18:19
Operator : SM/AJ
Sample : K3591-09
Misc : AR1660 MDL 25PPB
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019

Manual Integrations
APPROVED

Ankita
8/21/2019 12:58:08 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 01:05:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

