

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082019\
 Data File : P0059530.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2019 15:15
 Operator : SM/AJ
 Sample : K4396-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AOC-3-CENTER-5

Manual Integrations
 APPROVED

Ankita
 8/21/2019 1:59:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 16:23:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.139	3.504	859648	940416	24.553m	26.126m
2) SA Decachlor...	9.623	8.372	1047019	826108	19.681	20.488m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082019\
Data File : P0059530.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Aug 2019 15:15
Operator : SM/AJ
Sample : K4396-01
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AOC-3-CENTER-5


Manual Integrations
APPROVED

Ankita
8/21/2019 1:59:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 16:23:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.139 min
Delta R.T.: -0.010 min
Response: 859648
Conc: 24.55 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AOC-3-CENTER-5


Manual Integrations
APPROVED

Ankita
8/21/2019 1:59:29 PM




#1 Tetrachloro-m-xylene

R.T.: 3.504 min
Delta R.T.: -0.006 min
Response: 940416
Conc: 26.13 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.623 min
Delta R.T.: -0.018 min
Response: 1047019
Conc: 19.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.372 min
Delta R.T.: -0.011 min
Response: 826108
Conc: 20.49 ng/ml m