

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082119\
 Data File : P0059591.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Aug 2019 15:39
 Operator : SM/AJ
 Sample : K4396-06RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AOC-3-EAST-B-10RE

Manual Integrations
 APPROVED

Ankita
 8/22/2019 3:33:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 16:43:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.139	3.502	688842	653083	19.674	18.143
2) SA Decachlor...	9.624	8.371	514301	350029	9.667	8.681
Target Compounds						
26) L6 AR-1254-1	6.130	5.336	584480	577826	280.136	243.623
27) L6 AR-1254-2	6.348	5.482	1663061	448594	498.482	213.789 #
28) L6 AR-1254-3	6.722	5.894	1829237	1099317	515.502	326.934m#
29) L6 AR-1254-4	6.993	6.105	486529	380568	165.175	184.578
30) L6 AR-1254-5	7.409	6.512	776433	790141	251.078	254.476

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082119\
Data File : PO059591.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Aug 2019 15:39
Operator : SM/AJ
Sample : K4396-06RE
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AOC-3-EAST-B-10RE


Manual Integrations
APPROVED

Ankita
8/22/2019 3:33:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 16:43:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.139 min
Delta R.T.: -0.010 min
Response: 688842
Conc: 19.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AOC-3-EAST-B-10RE


Manual Integrations
APPROVED

Ankita
8/22/2019 3:33:19 PM




#1 Tetrachloro-m-xylene

R.T.: 3.502 min
Delta R.T.: -0.008 min
Response: 653083
Conc: 18.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.624 min
Delta R.T.: -0.016 min
Response: 514301
Conc: 9.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.371 min
Delta R.T.: -0.013 min
Response: 350029
Conc: 8.68 ng/ml




#26 AR-1254-1

R.T.: 6.130 min
Delta R.T.: -0.009 min
Response: 584480
Conc: 280.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AOC-3-EAST-B-10RE


Manual Integrations
APPROVED

Ankita
8/22/2019 3:33:19 PM




#26 AR-1254-1

R.T.: 5.336 min
Delta R.T.: -0.010 min
Response: 577826
Conc: 243.62 ng/ml




#27 AR-1254-2

R.T.: 6.348 min
Delta R.T.: -0.007 min
Response: 1663061
Conc: 498.48 ng/ml



#27 AR-1254-2

R.T.: 5.482 min
Delta R.T.: -0.010 min
Response: 448594
Conc: 213.79 ng/ml




#28 AR-1254-3

R.T.: 6.722 min
Delta R.T.: 0.002 min
Response: 1829237
Conc: 515.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AOC-3-EAST-B-10RE


Manual Integrations
APPROVED

Ankita
8/22/2019 3:33:19 PM




#28 AR-1254-3

R.T.: 5.894 min
Delta R.T.: 0.005 min
Response: 1099317
Conc: 326.93 ng/ml m




#29 AR-1254-4

R.T.: 6.993 min
Delta R.T.: -0.011 min
Response: 486529
Conc: 165.17 ng/ml



#29 AR-1254-4

R.T.: 6.105 min
Delta R.T.: -0.010 min
Response: 380568
Conc: 184.58 ng/ml




#30 AR-1254-5

R.T.: 7.409 min
Delta R.T.: -0.011 min
Response: 776433
Conc: 251.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AOC-3-EAST-B-10RE

Manual Integrations
APPROVED

Ankita
8/22/2019 3:33:19 PM



#30 AR-1254-5

R.T.: 6.512 min
Delta R.T.: -0.012 min
Response: 790141
Conc: 254.48 ng/ml