

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082119\
 Data File : P0059613.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Aug 2019 22:22
 Operator : SM/AJ
 Sample : K4448-10 5X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HOPPER-2-082019-75

Manual Integrations
 APPROVED

Ankita
 8/22/2019 3:34:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 04:01:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.130	3.500	198600	187008	5.672	5.195
2) SA Decachlor...	9.611	8.365	220103	177803	4.137	4.410m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082119\
Data File : P0059613.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Aug 2019 22:22
Operator : SM/AJ
Sample : K4448-10 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HOPPER-2-082019-75


Manual Integrations
APPROVED

Ankita
8/22/2019 3:34:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 04:01:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.130 min
Delta R.T.: -0.019 min
Response: 198600
Conc: 5.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
HOPPER-2-082019-75


Manual Integrations
APPROVED

Ankita
8/22/2019 3:34:22 PM




#1 Tetrachloro-m-xylene

R.T.: 3.500 min
Delta R.T.: -0.010 min
Response: 187008
Conc: 5.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.611 min
Delta R.T.: -0.029 min
Response: 220103
Conc: 4.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.365 min
Delta R.T.: -0.018 min
Response: 177803
Conc: 4.41 ng/ml m