

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082419\
 Data File : P0059828.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Aug 2019 2:41
 Operator : SM/AJ
 Sample : K4510-08
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-43-COMP

Manual Integrations
 APPROVED

Ankita
 8/26/2019 3:09:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 03:40:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.118	3.494	903476	851587	25.804m	23.658
2) SA Decachlor...	9.585	8.353	1081364	860640	20.326	21.345

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082419\
Data File : PO059828.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Aug 2019 2:41
Operator : SM/AJ
Sample : K4510-08
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-43-COMP


Manual Integrations
APPROVED

Ankita
8/26/2019 3:09:17 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 03:40:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.118 min
Delta R.T.: -0.031 min
Response: 903476
Conc: 25.80 ng/ml m

Instrument :
ECD_O
ClientSampleId :
SB-43-COMP


Manual Integrations
APPROVED

Ankita
8/26/2019 3:09:17 PM




#1 Tetrachloro-m-xylene

R.T.: 3.494 min
Delta R.T.: -0.016 min
Response: 851587
Conc: 23.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.585 min
Delta R.T.: -0.018 min
Response: 1081364
Conc: 20.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.353 min
Delta R.T.: -0.031 min
Response: 860640
Conc: 21.34 ng/ml