

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091523\
 Data File : P0098076.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2023 20:41
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/18/2023
 Supervised By :Ankita Jodhani 09/18/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 00:59:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.436	3.618	167.2E6	50758211	41.977	53.001 #
2)	SA Decachlor...	10.276	8.647	89938106	43697224	48.135m	51.145
Target Compounds							
3)	L1 AR-1016-1	5.613	4.707	46706490	17141170	432.547	532.930
4)	L1 AR-1016-2	5.636	4.725	68680010	23806676	424.401	524.081
5)	L1 AR-1016-3	5.698	4.902	41850157	12803904	420.295	525.896 #
6)	L1 AR-1016-4	5.797	4.945	32011725	10679850	405.957	508.347 #
7)	L1 AR-1016-5	6.095	5.159	33519034	14325207	408.298	515.436 #
31)	L7 AR-1260-1	7.231	6.196	59386454	26189895	420.229	492.248
32)	L7 AR-1260-2	7.491	6.385	64912027	31983876	429.437	510.768
33)	L7 AR-1260-3	7.854	6.539	48626085	28123571	429.259	487.710m
34)	L7 AR-1260-4	8.083	7.013	52760471	22927545	443.831	488.744
35)	L7 AR-1260-5	8.409	7.255	96274790	51417278	470.279	518.707

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091523\
Data File : PO098076.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2023 20:41
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/18/2023
Supervised By :Ankita Jodhani 09/18/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 16 00:59:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

